


Práctica de IMRT

Dr. Eduardo Francisco Larrinaga Cortina



QuickLinks Search Patient

- My
- Tasks ar
- Available
- Group By:
- Administration
- DICOM
- EMR
- Imaging
- Treatment Management
- Treatment Planning
- Quick Reports
- Customize...

Varianinstaller (INC...)

10/26/2023 - 10/26/2023

Due Date

Import Export

No records found

Launch (0) Same Window Separate Window

Patient List

Query List:

Available Completed

Your recent query list is empty. To select or create a new query, use the "Edit Queries" option in the Query Selection DropDown.

Launch(0) Currently showing 0/0 in the Patient List.

Schedule

New View Edit View Remove View

Today Work Week New Appointment New Task Check In Complete

	10/23/2023	10/24/2023	10/25/2023	10/26/2023	10/27/2023
8 am					
9 ⁰⁰					
10 ⁰⁰					
11 ⁰⁰					

Name	Type	Shared
DICOM IPA	DICOM Media File Import Filter	yes

Manage...

D:\Varian\Data\DICOM Refresh Scan subdirectories

Filename	Size	Date Modified
[Directory contains no relevant files]		



Browse For Folder

Select Import Directory

- Downloads
- Music
- Pictures
- Videos
- SYSTEM (C:)
- DATA (D:)
 - dicom
 - MSSQL
 - Program Files
 - Varian
 - Data


Make New Folder OK Cancel

- [ANON42620 - Breast_Intact2](#)
- [3metP4 - EW/HwZUq7y;mCvc2L\)gF4\('MN;t ...](#)
- [ROQS-24 - ROQS-24 SBRT LUNG](#)
- [2583839 - Sacky, Tom](#)

D:\dicom

... Refresh Scan subdirectories

Filename	Size	Date Modified
dataset mama\2.16.840.1.114362.1.6.6.6.16802.10793641352.448053110.815.2.dcm	513 KB	10/26/2023 7:36 AM
dataset mama\2.16.840.1.114362.1.6.6.6.16802.10793641352.448053110.829.5.dcm	513 KB	10/26/2023 7:36 AM
dataset mama\2.16.840.1.114362.1.6.6.6.16802.10793641352.448053110.836.6.dcm	513 KB	10/26/2023 7:36 AM
dataset mama\2.16.840.1.114362.1.6.6.6.16802.10793641352.448053110.842.7.dcm	513 KB	10/26/2023 7:36 AM
dataset mama\2.16.840.1.114362.1.6.6.6.16802.10793641352.448053110.847.8.dcm	513 KB	10/26/2023 7:36 AM
dataset mama\2.16.840.1.114362.1.6.6.6.16802.10793641352.448053110.853.9.dcm	513 KB	10/26/2023 7:36 AM



- [Back to Patients](#)
- 3metP4 - EWHwZUq7y;mCyc2LqF...
 - CT #134100 HFS
 - RTDOSE #134106
 - RTPLAN #134101
 - RTSTRUCT #134105
- Change Object ID...

D:\dicom

Refresh Scan subdirectories

Filename	Size	Date Modified
dataset mama\2.16.840.1.114362.1.6.6.6.16802.10793641352.448053110.815.2.dcm	513 KB	10/26/2023 7:36 AM
dataset mama\2.16.840.1.114362.1.6.6.6.16802.10793641352.448053110.829.5.dcm	513 KB	10/26/2023 7:36 AM
dataset mama\2.16.840.1.114362.1.6.6.6.16802.10793641352.448053110.836.6.dcm	513 KB	10/26/2023 7:36 AM
dataset mama\2.16.840.1.114362.1.6.6.6.16802.10793641352.448053110.842.7.dcm	513 KB	10/26/2023 7:36 AM
dataset mama\2.16.840.1.114362.1.6.6.6.16802.10793641352.448053110.847.8.dcm	513 KB	10/26/2023 7:36 AM
dataset mama\2.16.840.1.114362.1.6.6.6.16802.10793641352.448053110.853.9.dcm	513 KB	10/26/2023 7:36 AM



Filter Selection


- 3metP4 - EWHwZUq7y;mCyc2L)qF...
- CT #134100 HFS
- RTDOSE #134106
- RTPLAN #134101
- RTSTRUCT #134105

Change Object ID...

Patient Explorer

New Patient

Personal Demographics Address More


Last Name: 3Met
First Name: Cerebro
Middle Name:
ID1: 3metP4
Universal ID:
SSN:
Birth Date:
Age:
Sex: Unknown Marital:
E-mail Address:
Home Phone:
Work Phone:
Mobile Phone:
Mobile Phone Provider:
Other Phone:
Pager Number:
Fax Number:


Patient ID

Hospital: INCA Show history

Patient ID	Value	Entered On	Expiry Date
Global IDs			
ID 1	3metP4	10/26/2023	Active
ID 2			

Oncologists and Staff Referring Physician


 Edit

Primary	ID	Display Name
---------	----	--------------

Hospital

Name: INCA
Status: New Patient (NP)

Patient Access

 Edit

Primary	Department	Hospital
<input checked="" type="checkbox"/>	Radioterapia	INCA

OK Cancel

Summary

New Patient...
Browse Patients...

3metP4 - EWHwZUq7y;mCyc2L)qF...

- CT #134100 HFS
- RTDOSE #134106
- RTPLAN #134101
- RTSTRUCT #134105

Change Object ID...

Patient Selection - Please select an existing patient or create a new one.

Existing Patient(s)

Patient ID	First Name	Last Name	Patient Type
3metP4	Cerebr	3Met	New Patient

Import Export Log

Object / Date	Level	Category	Log Message
RTDOSE #134106 / Plan - ... 10/26/2023 7:49:59 AM	Warning	Object Conversion	DVH import is not supported - histogram(s) will be skipped
RTPLAN #134101 / HyperA... 10/26/2023 7:49:59 AM	Error	Object Connection	No matching machine found (Machine Id: 'Edge1' Manufacturer Name: 'Varian Medical Systems' Model: 'TDS' Serial Number: '')
10/26/2023 7:49:59 AM	Error	DICOM	Could not convert DICOM stream: SOP Instance UID: 1.2.246.352.999.474179248226601654183209545562742667718923301874 [Plan]
RT Dose: Plan - absolute (1... 10/26/2023 7:53:39 AM	Error	Object Connection	Not possible to connect RT Dose Plan to Plan, because referenced Plan '1.2.246.352.999.47417924822660165418320954556274266771
10/26/2023 7:53:39 AM	Error	Object Connection	RT Dose cannot be imported, plan not found
10/26/2023 7:53:39 AM	Error	Object Connection	Copy of RT Dose Plan failed
10/26/2023 7:53:41 AM	Error	Object Connection	Connection failed.

Group By Object

All
 Errors and Warnings
 Errors

All
 Reading Logs
 Connection Logs

Errors or warnings have occurred during the conversion process. Would you like to continue?


New Patient...

Browse Patients...

- 3metP4
 - DICOM
 - Study 5
 - Series Series (CT)
 - Image 0 ✓
 - Image 1 ✓
 - Image 2 ✓
 - Image 3 ✓
 - Image 4 ✓
 - Image 5 ✓
 - Image 6 ✓
 - Image 7 ✓
 - Image 8 ✓
 - Image 9 ✓
 - Image 10 ✓
 - Image 11 ✓
 - Image 12 ✓
 - Image 13 ✓
 - Image 14 ✓
 - Image 15 ✓
 - Image 16 ✓
 - Image 17 ✓
 - Image 18 ✓
 - Image 19 ✓
 - Image 20 ✓
 - Image 21 ✓
 - Image 22 ✓
 - Image 23 ✓
 - Image 24 ✓
 - Image 25 ✓
 - Image 26 ✓
 - Image 27 ✓
 - Image 28 ✓
 - Image 29 ✓
 - Image 30 ✓
 - Image 31 ✓
 - Image 32 ✓
 - Image 33 ✓
 - Image 34 ✓

Modality	Succeeded	Failed
Plan	0	1
CT	254	0
RT Dose	0	1
Structure Set	1	0
Volume Image	1	0
Total	256	2

View Detailed Log ...



QuickLinks Search Patient

My

- Favorites
- Contouring
- Categories
- Administration
- DICOM
- EMR
- Imaging
- Treatment Management
- Treatment Planning
 - Beam Configuration
 - Brachytherapy 2D Entry
 - Brachytherapy Planning
 - Contouring
 - External Beam Planning
 - IRREG Planning
 - Plan Evaluation
 - Selection
- Quick Reports
- Customize...

Varianinstaller (INC...)

10/26/2023 - 10/26/2023

Due Date

No records found

Patient List

Query List: Available Completed

Your recent query list is empty. To select or create a new query, use the "Edit Queries" option in the Query Selection DropDown.

Launch(0) Currently showing 0/0 in the Patient List.

Schedule

New View Edit View Remove View





Today Work Week New Appointment New Task Check In Complete

	10/23/2023	10/24/2023	10/25/2023	10/26/2023	10/27/2023
8 am					
9 ⁰⁰					
10 ⁰⁰					
11 ⁰⁰					

Launch (0) Same Window Separate Window





Transversal - SRS3x_120815 - 9/1/2017 5:08 PM

Frontal - SRS3x_120815 - 9/1/2017 5:08 PM



Transversal - SRS3x_120815 - 9/1/2017 5:08 PM

Frontal - SRS3x_120815 - 9/1/2017 5:08 PM







- SRS3x_120815
- 1GTV_RF
- 1PTV_RF_21
- 2GTV_RC
- 2PTV_RC_21
- 3GTV_LC
- 3PTV_LC_21
- ALIGNRT_OUTER
- BODY
- BRAIN
- BRAIN_PTV
- Brainstem
- Center
- Chiasm
- Coch_L
- Coch_R
- Cord
- Encompass
- Eye_L
- Eye_R
- HDoverride
- Lens_L
- Lens_R
- OptNrv_L
- OptNrv_R
- prior_LP_MD
- prior_PTV_MD
- prior_R_Front_MD
- prior_LP_Fuse
- PRIOR_RC_FUSE
- prior_RP_FUSE
- Z GI-1
- Z GI-2
- Z GI-3
- Z PTV1&2

Standard, HFS
Z: 34.65 cm






Transversal - SRS3x_120815 - 9/1/2017 5:08 PM

Frontal - SRS3x_120815 - 9/1/2017 5:08 PM



Transversal - SRS3x_120815 - 9/1/2017 5:08 PM

Frontal - SRS3x_120815 - 9/1/2017 5:08 PM



- SRS3x_120815
- 1GTV_RF
- 1PTV_RF_21
- 2GTV_RC
- 2PTV_RC_21
- 3GTV_IC
- 3PTV_IC_21
- ALIGNRT_OUTER
- BODY
- BRAIN
- BRAIN_PTV
- Brainstem
- Center
- Chiasm
- Coch_L
- Coch_R
- Cord
- Encompass
- Eye_L
- Eye_R
- HDoverride
- Lens_L
- Lens_R
- OptNrv_L
- OptNrv_R
- prior_LP_MD
- prior_PTV_MD
- prior_R_Front_MD
- prior_LP_Fuse
- PRIOR_RC_FUSE
- prior_RP_FUSE
- Z GI-1
- Z GI-2
- Z GI-3
- Z PTV1&2

Standard, HFS
Z: 34.65 cm

- PS Plan Studies
- CA Contouring Accuracy




Measure quality, study variation, and share best practices with a global community.

ProKnow® Plan Studies

What is it?

How does it work?



SEARCH MEDICAL AFFAIRS...

Halcyon Ethos TrueBeam & Edge (HD MLC) ProBeam

Cranial Head / Neck Thorax / Breast Abdomen Pelvis Extremity / Other Index

Case Studies

Cranial

- Whole Brain 30Gy
- Whole brain w/ hippocampalsparing 30Gy (NRG-CC001) [RP (HSWBv2)]
- Whole brain moderate hippocampalsparing EnhancedCoverage30Gy[RP]
- Whole brain limited hippocampal sparing EnhancedCoverage20Gy[RP]
- Whole brain w/ hippocampal sparing SIB 51/30Gy (2015 AAMD / SNC)
- CNS GBM 60/46Gy (RTOG 0825)
- Small single brain metastasis 24Gy
- Multimet (3) brain metastasis 21Gy
- Multimet (5) brain metastasis 21/18Gy



- Bilat head & neck 63/60/57/54Gy (2023 AAMD Plan Study-Phase 1)
- Bilateral Head&Neck 70/63/56Gy (2017 GADS/ProKnow Plan Study)
- Bilateral Head&Neck 70/63/56Gy (2013 Pinnacle/ROR Plan Challenge)

Feedback

Halcyon | Ethos | TrueBeam & Edge (HD MLC) | ProBeam

Cranial | Head / Neck | Thorax / Breast | Abdomen | Pelvis | Extremity / Other | Index


Multimet (3) brain metastasis

Rx: 21Gy/1fx

THREE BRAIN METS ON DIFFERENT PLANES



VMAT SX1 | VMAT SX2



- Beams >
- DVH >
- Simple DVH Metrics >
- Technical Plan Comments >

Feedback


Whole brain w/ hippocampalsparing 30Gy (NRG-CC001) [RP (HSWBv2)]

Rx: 30Gy/10fx

NRG-CC001



VMAT SX1 | VMAT SX2



- Beams >
 - DVH >
 - Scorecard Metrics >
 - Technical Plan Comments >
- Four (4) arc VMAT technique was chosen and 4 almost complete arcs were used. This plan was made on Halcyon with the MLC in SX2 mode and AcurosXB v17.0. Plan created using the RapidPlan HSWBv2 model seen below with convergence mode: extended and MU objective enabled with Minimum MU set to 1600. Full details in clinical description document packaged with the RapidPlan model or at the PDF link below.
- Maximal hippocampal sparing, beyond what is requested by the NRG-CC001 protocol, is achieved here. This HSWBv2 model (released 2022) was derived from the HSWBv1 (released 2016) and utilized a similar recursive model creation process. This time all trainingset cases are made from an initial HSWBv2 model but the trainingset cases are all Halcyon SX2 plans calculated with AcurosXB v17. Additional structures and metrics were added to the dosimetric scorecard which guided the model creation process and helped further reduce hippocampal dose, improve brain homogeneity, dose conformity and lower doses in the patient's face. This HSWBv2 model offers unparalleled plan quality on the Halcyon delivery platform in an easy to use single click solution.
- Physician Comments >


Halcyon Ethos TrueBeam & Edge (HD MLC) ProBeam

Cranial Head / Neck Thorax / Breast Abdomen Pelvis Extremity / Other Index

Craniospinal – CNS

Rx: 23.4Gy/13fx

10 arc, 5 isocenter treatment delivery



Beams >

DVH >

Simple DVH Metrics >

Technical Plan Comments >

3rd party software plan report
[PDF](#)

DICOM patient export
[ZIP](#)

Any reference to a "plan study" are simply what the organizers call each case and may not be a "study" in the FDA sense as they may not have been published in a peer reviewed journal.
Varian does not provide medical advice and these are illustrative examples only.
Leading plans by expert planner. Your results may vary.

FOR EDUCATIONAL AND SCIENTIFIC EXCHANGE ONLY – NOT FOR SALES OR PROMOTIONAL USE.

Feedback

Transversal - HCSWB V2.0 - 10/31/2014 1:43 PM

Frontal - HCSWB V2.0 - 10/31/2014 1:43 PM

The interface displays a CT scan of a head with various anatomical structures contoured. The main window shows a Transversal view. To the right, there are smaller windows for Frontal and Sagittal views. A 3D view is shown at the bottom right. A legend on the left lists the contoured structures.

- HCSWB V2.0
- _Brain&BODY
- _Brainstem#Hi
- _Eyes&Body
- BODY
- Brain
- Brainstem
- Chiasm
- Cord
- CouchInterior
- CouchSurface
- Eyes
- Hippo+05
- Hippocampus
- Leye
- LLacrimial
- LLens
- Loptic
- NS_Artifact
- PTV_WB
- PTV_WBopt
- Reye
- RLacrimial
- RLens
- Roptic
- User Origin


Standard, HFS
Z: -1.00 cm

Y: -1.24 cm

X: -0.04 cm

The interface displays a CT scan of a head with various anatomical structures contoured. The main window shows a Transversal view with a central crosshair and a vertical red line. The left sidebar lists the following contours: HCSWB V2.0, _Brain&BODY, _Brainstem#Hi, _Eyes&Body, BODY, Brain, Brainstem, Chiasm, Cord, CouchInterior, CouchSurface, Eyes, Hippo+05, Hippocampus, Leye, LLacriml, LLens, Loptic, NS_Artifact, PTV_WB, PTV_WBopt, Reye, RLacriml, RLens, Roptic, and User Origin. The bottom status bar shows 'Standard, HFS Z: -1.00 cm'. The top right corner shows 'Frontal - HCSWB V2.0 - 10/31/2014 1:43 PM'. The middle right corner shows 'Sagittal - HCSWB V2.0 - 10/31/2014 1:43 PM'. The bottom right corner shows '3D - HCSWB V2.0 - 10/31/2014 1:43 PM'. The bottom right corner also shows a 3D view of the head with a red plane and a green crosshair. The bottom right corner also shows a 3D view of the head with a red plane and a green crosshair. The bottom right corner also shows a 3D view of the head with a red plane and a green crosshair.


Transversal - RTstruct-Va - 5/4/2011 4:34 PM



- C-Brain
- C-Brainstem
- C-Chiasm
- C-Cord
- C-Eye_Rt
- C-Eye_Lt
- C-Iam+Cochlr_Lt
- C-Iam+Cochlr_Rt
- C-Larynx
- C-Lens- Lt
- C-Lens- Rt
- C-Optic Nerve_Lt
- C-Optic Nerve_Rt
- C-Oral cavity
- CORD+1
- Couch
- CouchInterior
- CouchSurface
- C-Parotid-Lt
- C-Parotid-Rt
- C-Skin
- PH2_CTV
- PTV_Cranium
- PTV_Spine
- S-Heart
- S-Kidney_Lt
- S-Kidney_Rt
- S-Liver
- S-Lung_Lt
- S-Lung_Rt
- S-Oesophagus
- S-Thyroid
- S-Total Kidney
- S-Total Lung
- User Origin

Standard, HFS
Z: 4.50 cm


Frontal - RTstruct-Va - 5/4/2011 4:34 PM



Sagittal - RTstruct-Va - 5/4/2011 4:34 PM





3D - RTstruct-Va - 5/4/2011 4:34 PM



Transversal - RTstruct-Va - 5/4/2011 4:34 PM

Frontal - RTstruct-Va - 5/4/2011 4:34 PM



- C-Brain
- C-Brainstem
- C-Chiasm
- C-Cord
- C-Eye_Rt
- C-Eye_Lt
- C-Iam+Cochlr_Lt
- C-Iam+Cochlr_Rt
- C-Larynx
- C-Lens- Lt
- C-Lens- Rt
- C-Optic Nerve_Lt
- C-Optic Nerve_Rt
- C-Oral cavity
- CORD+1
- Couch
- CouchInterior
- CouchSurface
- C-Parotid-Lt
- C-Parotid-Rt
- C-Skin
- PH2_CTV
- PTV_Cranium
- PTV_Spine
- S-Heart
- S-Kidney_Lt
- S-Kidney_Rt
- S-Liver
- S-Lung_Lt
- S-Lung_Rt
- S-Oesophagus
- S-Thyroid
- S-Total Kidney
- S-Total Lung
- User Origin

Standard, HFS
Z: -26.50 cm

most caudal extent of the crescentic-shaped floor of the temporal horn and continued postero-cranially along the medial edge of the temporal horn. The medial border of the hippocampus was demarcated by the edge of the T1-hypointensity up to the ambient cistern. The uncus recess of the temporal horn served to distinguish the hippocampus from the gray matter of the amygdala, lying anterior and superior to the hippocampus. The postero-cranial extent of the hip-

sampled the CT image resolution to 256×256 pixels per slice, the slice width was maintained at 2.5 mm for the entire CT image volume set. Plans were optimized such that 96% of the whole-brain PTV received the prescription dose of 30 Gy in 10 fractions. Helical tomotherapy plan parameters consisted of a 1.05-cm field width, 0.215 pitch, and 3.0 modulation factor, based on dosimetric results from a prior helical tomotherapy planning study (9).

Table 1. Clinical criteria and inverse planning algorithm constraints for helical tomotherapy and LINAC-based IMRT planning

Structure	Helical tomotherapy plan criteria	Penalty	Importance	LINAC-based IMRT plan criteria	Penalty
Whole-brain PTV	Max dose: 30 Gy 30 Gy to $\geq 96\%$	100	200	Max dose: 34 Gy Min dose: 32 Gy	100
Hippocampus	Max dose: 6 Gy 3 Gy to $\leq 20\%$	100 20	500	Max dose: 11 Gy 9 Gy to $\leq 40\%$	5 10
Hippocampal avoidance volume	Max dose: 30 Gy 20 Gy to $\leq 20\%$	1 10	5	N/A	N/A
Eyes*	Max dose: 8 Gy 5 Gy to $\leq 20\%$	10 10	20	Max dose: 7 Gy	5
Lenses*	Max dose: 3 Gy 2 Gy to $\leq 20\%$	20 10	20	Max dose: 5 Gy	5

Abbreviations: IMRT = intensity-modulated radiotherapy; LINAC = linear accelerator; Max = maximum; N/A, not applicable; PTV = planned target volume.

* Eyes and lenses were directionally blocked during helical tomotherapy planning.

Table 2. Starting plan configuration in x, y, z coordinate system for non-coplanar treatment planning, which follows a basic template provided by the Plan Geometry



doi:10.1016/j.ijrobp.2010.01.039

PHYSICS CONTRIBUTION

HIPPOCAMPAL-SPARING WHOLE-BRAIN RADIOTHERAPY: A “HOW-TO” TECHNIQUE USING HELICAL TOMOTHERAPY AND LINEAR ACCELERATOR-BASED INTENSITY-MODULATED RADIOTHERAPY

VINAI GONDI, M.D.,* RANJINI TOLAKANAHALLI, M.S.,[†] MINESH P. MEHTA, M.D.,*
DINESH TEWATIA, M.S.,*[†] HOWARD ROWLEY, M.D.,[‡] JOHN S. KUO, M.D., PH.D.,*[§]
DEEPAK KHUNTIA, M.D.,* AND WOLFGANG A. TOMÉ, PH.D.*[†]

Departments of *Human Oncology, [†]Medical Physics, [‡]Neuroradiology, and [§]Neurological Surgery, University of Wisconsin
Comprehensive Cancer Center, Madison, WI






HCSWB

- ANON96873
 - Series
 - HCSWB V2.0

HCSWB V2.0

- Registered Images
 - HCSWB V2.0
 - _Brain&BODY
 - _Brainstem#Hi
 - _Eyes&Body
 - BODY
 - Brain
 - Brainstem
 - Chiasm
 - Cord
 - CouchInterior
 - CouchSurface
 - Eyes
 - Hippo+05
 - Hippocampus
 - Leye
 - LLacrimal
 - LLens
 - Loptic
 - NS_Artifact
 - PTV_WB
 - PTV_WBopt
 - Reye
 - RLacrimal
 - RLens



Fields Dose Field Alignments Plan Objectives Optimization Objectives Dose Statistics Reference Points Calculation Models Plan Sum



Group	Field ID	Technique	Machine/Energy	MLC	Field Weight	Scale	GantryRtn [deg]	CollRtn [deg]	PatientSupportAngle [deg]	X [cm]	Y [cm]	Z [cm]	Calculated SSD [cm]	Meterset

New Course...

- New Clinical Protocol Reference...
- New 4D Image...
- New Structure Set...
- New Structures from Template...
- New Structure...
- New Isocenter Marker...
- New Marker...
- New Bolus...
- New Couch Structures...
- New Plan from Template... F8
- New Plan... Ctrl+F8
- New HyperArc Plan...
- New Plan to Plan Sum...
- New Plan Sum...
- New Field... F9
- New Setup Field...
- New Opposing Field F10
- New Field in Field
- New Imaging Setup...
- New Field to New Isocenter Group...
- New Block...
- New MLC...
- New DRR...
- New Irregular Surface Compensator...
- New Reference Point and Location...
- New Location for Existing Reference Point...

HCSWB V2.0

- Registered
- HCSWB V2.0
 - _Brain
 - _Brain
 - _Eyes
 - BODY
 - Brain
 - Brain:
 - Chias
 - Cord
 - Couch
 - Couch
 - Eyes
 - Hippe
 - Hippe
 - Leye
 - LLacr
 - LLens
 - Lopti
 - NS_A
 - PTV_V
 - PTV_V
 - Reye
 - RLacrima
 - RLens





Group	Field ID	Technique	Machine/Energy	MLC	Field Weight	Scale	GantryRtn [deg]	CollRtn [deg]	PatientSupportAngle [deg]	X [cm]	Y [cm]	Z [cm]	Calculated SSD [cm]	Meterset

File Edit View Insert Planning Tools Window

- New Course...
- New Clinical Protocol Reference...
- New 4D Image...
- New Structure Set...
- New Structures from Template...
- New Structure...
- New Isocenter Marker...
- New Marker...
- New Bolus...
- New Couch Structures...
- New Plan from Template...
- New Plan... F8**
- New HyperArc Plan... Ctrl+F8
- New Plan to Plan Sum...
- New Plan Sum...
- New Field... F9
- New Setup Field...
- New Opposing Field F10
- New Field in Field
- New Imaging Setup...
- New Field to New Isocenter Group...
- New Block...
- New MLC...
- New DRR...
- New Irregular Surface Compensator...
- New Reference Point and Location...
- New Location for Existing Reference Point...

HCSWB V2.0

- Registered
- HCSWB V2.0
 - _Brain
 - _Brain
 - _Eyes
 - BODY
 - Brain
 - Brain:
 - Chias
 - Cord
 - Couch
 - Couch
 - Eyes
 - Hippe
 - Hippe
 - Leye
 - LLacr
 - LLens
 - Lopti
 - NS_A
 - PTV_V
 - PTV_V
 - Reye
 - RLacrima
 - RLens



Fields Dose Field Alignments Plan Objectives Optimization Objectives Dose Statistics Reference Points Calculation Models Plan Sum

Group	Field ID	Technique	Machine/Energy	MLC	Field Weight	Scale	GantryRtn [deg]	CollRtn [deg]	PatientSupportAngle [deg]	X [cm]	Y [cm]	Z [cm]	Calculated SSD [cm]	Meterset

HCSWB

- ANON96873
 - Series
 - HCSWB V2.0
 - Etapa1

HCSWB V2.0

- Registered Images
- HCSWB V2.0
 - _Brain&BODY
 - _Brainstem#Hi
 - _Eyes&Body
 - BODY
 - Brain
 - Brainstem
 - Chiasm
 - Cord
 - CouchInterior
 - CouchSurface
 - Eyes
 - Hippo+05
 - Hippocampus
 - Leye
 - LLacrimal
 - LLens
 - Loptic
 - NS_Artifact
 - PTV_WB
 - PTV_WBopt
 - Reye
 - RLacrimal
 - RLens

Transversal - HCSWB V2.0

3D DVH BEV Arc HCSWB V2.0 - Model View - HCSWB V2.0

Frontal - HCSWB V2.0

Z: -5.00

Y: 0.13 cm

X: -0.04 cm

Select Course

Select a course for the new plan from the list of available courses.

Available courses

- Etapa1

Course details

ID Etapa1

Intent

Status ACTIVE

New Course...

< Back Next > Cancel Help

Group	Field ID	Technique	Machine/Energy	MLC	Field Weight	Scale	GantryRtn [deg]	CollRtn [deg]	PatientSupportAngle [deg]	X [cm]	Y [cm]	Z [cm]	Calculated SSD [cm]	Meterset
-------	----------	-----------	----------------	-----	--------------	-------	-----------------	---------------	---------------------------	--------	--------	--------	---------------------	----------

HCSWB
ANON96873
Series
HCSWB V2.0
Etapa1

- HCSWB V2.0
- Registered Images
- HCSWB V2.0
 - _Brain&BODY
 - _Brainstem#Hi
 - _Eyes&Body
 - BODY
 - Brain
 - Brainstem
 - Chiasm
 - Cord
 - CouchInterior
 - CouchSurface
 - Eyes
 - Hippo+05
 - Hippocampus
 - Leye
 - LLacrimal
 - LLens
 - Loptic
 - NS_Artifact
 - PTV_WB
 - PTV_WBopt
 - Reye
 - RLacrimal
 - RLens

Transversal - HCSWB V2.0
Frontal - HCSWB V2.0
3D DVH BEV Arc HCSWB V2.0 - Model View - HCSWB V2.0

Plan Details


Course	ID: Etapa1	Dose	Dose per Fraction: 300.000 Gy
			Number of Fractions: 10
			Total Dose: 3000.000 Gy
			Treatment Percentage: 100.00 %
Plan	ID: Plan1		
	Name: HCSWB-VMAT		
	Intent: Curative		

< Back Next > Cancel Help

Group	Field ID	Technique	Machine/Energy	MLC	Field Weight	Scale	GantryRtn [deg]	CollRtn [deg]	PatientSupportAngle [deg]	X [cm]	Y [cm]	Z [cm]	Calculated SSD [cm]	Meterset
-------	----------	-----------	----------------	-----	--------------	-------	-----------------	---------------	---------------------------	--------	--------	--------	---------------------	----------

HCSWB
ANON96873
Series
HCSWB V2.0
Etapa1

- HCSWB V2.0
- Registered Images
- HCSWB V2.0
 - _Brain&BODY
 - _Brainstem#Hi
 - _Eyes&Body
 - BODY
 - Brain
 - Brainstem
 - Chiasm
 - Cord
 - CouchInterior
 - CouchSurface
 - Eyes
 - Hippo+05
 - Hippocampus
 - Leye
 - LLacrimal
 - LLens
 - Loptic
 - NS_Artifact
 - PTV_WB
 - PTV_WBopt
 - Reye
 - RLacrimal
 - RLens



Select Target Volume

No target volume

Available Volumes

Volume	Type
PTV_WB	PTV
PTV_WBopt	PTV
_Brain&BODY	Control
_Brainstem#Hi	Control
_Eyes&Body	Control
BODY	BODY
Brain	Organ
Brainstem	Organ
Chiasm	Organ
Cord	Organ

Volume Details

ID: PTV_WBopt
Name: PTV_WBopt

New Target Volume...

< Back Next > Cancel Help

Group	Field ID	Technique	Machine/Energy	MLC	Field Weight	Scale	GantryRtn [deg]	CollRtn [deg]	PatientSupportAngle [deg]	X [cm]	Y [cm]	Z [cm]	Calculated SSD [cm]	Meterset
-------	----------	-----------	----------------	-----	--------------	-------	-----------------	---------------	---------------------------	--------	--------	--------	---------------------	----------

- HCSWB
 - ANON96873
 - Series
 - HCSWB V2.0
 - Etapa1
- _Brain&BODY
 - _Brainstem#Hi
 - _Eyes&Body
 - BODY
 - Brain
 - Brainstem
 - Chiasm
 - Cord
 - CouchInterior
 - CouchSurface
 - Eyes
 - Hippo+05
 - Hippocampus
 - Leye
 - LLacrimial
 - LLens
 - Loptic
 - NS_Artifact
 - PTV_WB
 - PTV_WBopt
 - Reye
 - RLacrimial
 - RLens
 - Roptic
- Reference Points
 - PTV WBopt

Transversal - HCSWB V2.0

3D DVH BEV Arc HCSWB V2.0 - Model View - HCSWB V2.0

Select Primary Reference Point

Select a primary reference point for the new p

Available Reference Points for PTV_WBopt

Frontal - HCSWB V2.0

Z: -5.00

Y: 0.13 cm

X: -0.04 cm

New Reference Point...

Cancel Help

Reference Point Properties

General History Debug

ID

PTV_WBopt

Name

Connection

Not connected to any plan

Patient Volume

PTV_WBopt [PTV] New...

Dose Limits

Total Dose Limit 30.000 Gy

Daily Dose Limit 3.000 Gy

Session Dose Limit 3.000 Gy

OK Cancel Help

Group	Field ID	Technique	Machine/Energy	MLC	Field Weight	Scale	GantryRtn [deg]	CollRtn [deg]	PatientSupportAngle [deg]	X [cm]	Y [cm]	Z [cm]	Calculated SSD [cm]	Meterset
-------	----------	-----------	----------------	-----	--------------	-------	-----------------	---------------	---------------------------	--------	--------	--------	---------------------	----------

HCSWB
ANON96873
Series
HCSWB V2.0
Etapa1

- ✓ _Brain&BODY
- ✓ _Brainstem#Hi
- ✓ _Eyes&Body
- ✓ BODY
- ✓ Brain
- ✓ Brainstem
- ✓ Chiasm
- ✓ Cord
- ✓ CouchInterior
- ✓ CouchSurface
- ✓ Eyes
- ✓ Hippo+05
- ✓ Hippocampus
- ✓ Leye
- ✓ LLacrimonal
- ✓ LLens
- ✓ Loptic
- ✓ NS_Artifact
- ✓ PTV_WB
- ✓ PTV_WBopt
- ✓ Reye
- ✓ RLacrimonal
- ✓ RLens
- ✓ Roptic
- Reference Points
 - PTV WBopt

Transversal - HCSWB V2.0

3D DVH BEV Arc HCSWB V2.0 - Model View - HCSWB V2.0

Frontal - HCSWB V2.0

Z: -5.00

Y: 0.13 cm

X: -0.04 cm

Select Treatment Machine

Available Treatment Machines

- Eclipse CAP

Selected Treatment Machine

ID Eclipse CAP

Name

Machine Model 2100C/D

Manufacturer Name Varian Medical Systems

MLC MLC120 (Millennium 120)

< Back Next > Cancel Help

Group	Field ID	Technique	Machine/Energy	MLC	Field Weight	Scale	GantryRtn [deg]	CollRtn [deg]	PatientSupportAngle [deg]	X [cm]	Y [cm]	Z [cm]	Calculated SSD [cm]	Meterset
-------	----------	-----------	----------------	-----	--------------	-------	-----------------	---------------	---------------------------	--------	--------	--------	---------------------	----------

- HCSWB
 - ANON96873
 - Series
 - HCSWB V2.0
 - Etapa1
- _Brain&BODY
 - _Brainstem#Hi
 - _Eyes&Body
 - BODY
 - Brain
 - Brainstem
 - Chiasm
 - Cord
 - CouchInterior
 - CouchSurface
 - Eyes
 - Hippo+05
 - Hippocampus
 - Leye
 - LLacrima
 - LLens
 - Loptic
 - NS_Artifact
 - PTV_WB
 - PTV_WBopt
 - Reye
 - RLacrima
 - RLens
 - Roptic
- Reference Points
 - PTV WBopt

Transversal - HCSWB V2.0

3D DVH BEV Arc HCSWB V2.0 - Model View - HCSWB V2.0

Frontal - HCSWB V2.0

Z: -5.00

Y: 0.13 cm

X: -0.04 cm

Select Patient Position

Patient Position During Treatment

- Head First-Supine
- Head First-Supine**
- Head First-Prone
- Feet First-Supine
- Feet First-Prone
- Head First-Decubitus Left
- Head First-Decubitus Right
- Feet First-Decubitus Left
- Feet First-Decubitus Right

Image

Study: ANON96873

Series: Series

Image ID: HCSWB V2.0

Creation Date: 10/31/2014

Creation Time: 1:43:16 PM

< Back Next > Cancel Help


Group	Field ID	Technique	Machine/Energy	MLC	Field Weight	Scale	GantryRtn [deg]	CollRtn [deg]	PatientSupportAngle [deg]	X [cm]	Y [cm]	Z [cm]	Calculated SSD [cm]	Meterset
-------	----------	-----------	----------------	-----	--------------	-------	-----------------	---------------	---------------------------	--------	--------	--------	---------------------	----------

HCSWB

- ANON96873
 - Series
 - HCSWB V2.0
 - Etapa1
 - Plan1

Plan1

- HCSWB V2.0
 - Registered Images
 - HCSWB V2.0
 - _Brain&BODY
 - _Brainstem#Hi
 - _Eyes&Body
 - BODY
 - Brain
 - Brainstem
 - Chiasm
 - Cord
 - CouchInterior
 - CouchSurface
 - Eyes
 - Hippo+05
 - Hippocampus
 - Leye
 - LLacrimal
 - LLens
 - Loptic
 - NS_Artifact
 - PTV_WB
 - PTV_WBopt
 - Reye
 - RLacrimal



Field Properties

Reference Image Setup Notes History Comment Debug

General Geometry Room Geometry Accessories Calculation

ID: Field 1

Name:

Machine: Eclipse CAP

Physical

Energy: 6X

Dose Rate: 300 MU/min

MU: MU


Planned Field Dose: Gy

SAD: 100.0 cm

Tolerance: T1

Time: min

OK Cancel Help







Group	Field ID	Technique	Machine/Energy	MLC	Field Weight	Scale	Field Y [cm]	Y1 [cm]	Y2 [cm]	X [cm]	Y [cm]	Z [cm]	Calculated SSD [cm]	MU	Ref. D [Gy]
1	Field 1	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	0.0	0.0	0.0	0.24	1.58	1.09	89.9		

HCSWB

- ANON96873
 - Series
 - HCSWB V2.0
 - Etapa1
 - Plan1

- Chiasm
- Cord
- CouchInterior
- CouchSurface
- Eyes
- Hippo+05
- Hippocampus
- Leye
- LLacrimonal
- LLens
- Loptic
- NS_Artifact
- PTV_WB
- PTV_WBopt
- Reye
- RLacrimonal
- RLens
- Roptic
- User Origin
- Reference Points
 - PTV_WBopt
- Dose
- Fields
 - Isocenter Group I
 - Field 1



Group	Field ID	Technique	Machine/Energy	MLC	Field Weight	Scale	Gantry Rtn [deg]	Coll Rtn [deg]	Couch Rtn [deg]	Wedge	Field X [cm]	X1 [cm]	X2 [cm]	Field Y [cm]	Y1 [cm]	Y2 [cm]	X [cm]	Y [cm]	Z [cm]	Calculated SSD [cm]	MU	Ref. D [Gy]
I	Field 1	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	0.0	0.0	0.0	None	10.0	-5.0	+5.0	10.0	-5.0	+5.0	0.24	1.58	1.09	89.9		




Arc Geometry Tool... Ctrl+F7

- HyperArc
- Optimization
- Dose Calculation
- Create Verification Plan...
- Create Partial Treatment Plan...
- Plan Normalization...
- Isodose Levels...
- Compensator Isolevels...
- Plan Uncertainty Parameters...
- Show Dose Volume Histogram View
- Create Plan Comparison DVH...
- Biological Evaluation...
- Verify MLC Leaf Positions...
- Field Weight... F3
- Change Treatment Units...
- DVH Based Plan Converter...
- Merge Subfields
- Edit Compensator...
- Field Order...
- Reference Point Organizer...
- Delta Couch Shift Editor...
- Templates and Clinical Protocols
- Enable Digitizer
- Check IHE-RO RT Compliance

HCSWB

- ANON96873
- Series
- HCSWB V2.0
- Etapa1
- Plan1

- Chiasm
- Cord
- CouchInte
- CouchSur
- Eyes
- Hippo+O3
- Hippocan
- Leye
- LLacrimal
- LLens
- Loptic
- NS_Artifa
- PTV_WB
- PTV_WBo
- Reye
- RLacrimal
- RLens
- Roptic
- User Origin
- Reference Points
- PTV_WBopt
- Dose
- Fields
- Isocenter Group 1
- Field 1





Group	Field ID	Technique	Machine/Energy	MLC	Field Weight	Scale	Gantry Rtn [deg]	Coll Rtn [deg]	Couch Rtn [deg]	Wedge	Field X [cm]	X1 [cm]	X2 [cm]	Field Y [cm]	Y1 [cm]	Y2 [cm]	X [cm]	Y [cm]	Z [cm]	Calculated SSD [cm]	MU	Ref. D [Gy]
I	Field 1	STATIC-1	Eclipse CAP - 6X		1.000	IEC61217	0.0	0.0	0.0	None	10.0	-5.0	+5.0	10.0	-5.0	+5.0	0.24	1.58	1.09	89.9		

HCSWB

- ANON96873
 - Series
 - HCSWB V2.0
 - Etapa1
 - Plan1

Plan1

- HCSWB V2.0
 - Registered Images
 - HCSWB V2.0
 - User Origin
 - Reference Points
 - PTV_WBopt
 - Dose
 - Fields
 - Isocenter Group 1
 - Field 1



Field Properties

Reference Image Setup Notes History Comment Debug

General Geometry Room Geometry Accessories Calculation

ID: Field 1

Scale: IEC61217

Technique: ARC

Setup: Isocentric

Use Gated

Gantry

Gantry Rtn: 181.0 deg Extended

Stop Angle: 179.0 deg Extended

Direction: None CW CCW

Collimator

Coll Rtn: 10.0 deg

Field X: 10.0 cm Field Y: 10.0 cm

X1: cm Y1: cm

X2: cm Y2: cm

Asym. X Asym. Y

Couch

Couch Rtn: 0.0 deg

Field

SFED: cm

SBD: cm

Isocenter

X: 0.24 cm

Y: 1.58 cm


Z: 1.09 cm

SSD

Calculated: 91.3 cm

Planned: cm

OK Cancel Apply Help



Use Default Models	Particle Type	Calculation Type	
Clear All Selections	Photon	Volume Dose	AcurosXB_156
		DVH Estimation	DVH Estimation
		Beam Angle Optimization	PGO_15161
		IMRT Optimization	PO_15606
		VMAT Optimization	PO_15606
		Irregular Surface Compensator	PO_15606
		Stereotactic Dose	

Algorithm	Calculation Options
Acuros External Beam (Version 15.6.06)	Edit
DVH Estimation Algorithm (Version 15.6.06)	Edit
Plan Geometry Optimizer (Version 15.1.61)	Edit
Photon Optimizer (Version 15.6.06)	Edit


HCSWB

- ANON96873
 - Series
 - HCSWB V2.0
 - Etapa1
 - Plan1


Plan1

- HCSWB V2.0
 - Registered Images
 - HCSWB V2.0
 - User Origin
 - Reference Points
 - PTV_WBopt
 - Dose
 - Fields
 - Isocenter Group 1
 - Field 1

Plan1 - Unapproved - Transversal - HCSWB V2.0



Plan1 - Unapproved - Frontal - HCSWB V2.0



Fit Collimator to Structure

Target field

ID: Field 1

Target structure

ID: PTV_WBopt

Margin

Circular [cm] 0.50

Elliptical

X1 [cm]

X2 [cm]

Y1 [cm]

Y2 [cm]

Coordinate system

BEV Collimator

Options


Use asymmetric X jaws

Use asymmetric Y jaws


Optimize collimator rotation (in BEV coordinate system only)

Fit Close

3D DVH BEV Arc Plan1 - Unapproved - BEV - SAD 100 cm - Field 1 - Arc viewing angle: 181.00 [deg] - HCSWB V2.0



- Unapproved - Sagittal - HCSWB V2.0



Use Default Models

Clear All Selections



Particle Type	Calculation Type	Calculation Model	Status
Photon	Volume Dose	AcurosXB_15606	OK
	DVH Estimation	DVH Estimation Algorithm [15.6.06]	OK
	Beam Angle Optimization	PGO_15161	No approved beam data (Eclipse CAP 6X OPENFIELD)
	IMRT Optimization	PO_15606	OK
	VMAT Optimization	PO_15606	OK
	Irregular Surface Compensator	PO_15606	Not supported for this plan
	Stereotactic Dose		

Algorithm	Calculation Options
Acuros External Beam (Version 15.6.06)	Edit
DVH Estimation Algorithm (Version 15.6.06)	Edit
Plan Geometry Optimizer (Version 15.1.61)	Edit
Photon Optimizer (Version 15.6.06)	Edit

HCSWB

- ANON96873
 - Series
 - HCSWB V2.0
 - Etapa1
 - Plan1
 - Cord
 - Couch
 - Couch
 - Eyes
 - Hippocampus
 - Hippocampus
 - Lens
 - Lens
 - LLacrimal
 - LLens
 - Optic
 - NS_Artery
 - PTV_Visual
 - PTV_Visual
 - Reye
 - RLacrimal
 - RLens
 - ROptic
 - User Origin
 - Reference Point
 - PTV_WBContour
 - Dose
 - Fields
 - Isocenter
 - Field 1
 - Field 2

- Copy Field
- Paste Field with Field Image(s)
- Delete
- Align Fields in Isocenter Group 1 by Field 1 to
- Fit Collimator to Structure
- Show Beam's Eye View
- Add Field Image...
- Show Field Dose
- Arrange as Multi-Isocenter Setup
- Reverse Arc Direction
- Move Viewing Planes to Isocenter/Entry Point
- Modify Isocenter Group...
- New Field...
- New Setup Field from Selected Treatment Field
- New Opposing Field
- New Field in Field
- New Block...
- New MLC...
- New DRR
- New Irregular Surface Compensator...
- New Imaging Setup...
- Import Optimal Fluence...
- Export Optimal Fluence...
- Change Treatment Units...
- Properties



Use Default Models	Particle Type	Calculation Type	Calculation Model	Status	Algorithm	Calculation Options
Clear All Selections	Photon	Volume Dose	AcurosXB_15606	OK	Acuros External Beam (Version 15.6.06)	Edit
		DVH Estimation	DVH Estimation Algorithm [15.6.06]	OK	DVH Estimation Algorithm (Version 15.6.06)	Edit
		Beam Angle Optimization	PGO_15161	No approved beam data (Eclipse CAP 6X OPENFIELD)	Plan Geometry Optimizer (Version 15.1.61)	Edit
		IMRT Optimization	PO_15606	OK	Photon Optimizer (Version 15.6.06)	Edit
		VMAT Optimization	PO_15606	OK		
		Irregular Surface Compensator	PO_15606	Not supported for this plan		
		Stereotactic Dose				

- Arc Geometry Tool... Ctrl+F7
- HyperArc
- Optimization**
 - Optimize... F7
 - Beam Angle Optimization...** Shift+F7
 - Conformal Optimization...
 - Biological Optimization...
- Dose Calculation
- Create Verification Plan...
- Create Partial Treatment Plan...
- Plan Normalization...
- Isodose Levels...
- Compensator Isolevels...
- Plan Uncertainty Parameters...
- Show Dose Volume Histogram View
- Create Plan Comparison DVH...
- Biological Evaluation...
- Verify MLC Leaf Positions...
- Field Weight... F3
- Change Treatment Units...
- DVH Based Plan Converter...
- Merge Subfields
- Edit Compensator...
- Field Order...
- Reference Point Organizer...
- Delta Couch Shift Editor...
- Templates and Clinical Protocols
- Enable Digitizer
- Check IHE-RO RT Compliance


Group	Field ID	Technique	Machine/Energy	MLC	Field Weight	Scale	Gantry Rtn [deg]	Coll Rtn [deg]	Couch Rtn [deg]	Wedge	Field X [cm]	X1 [cm]	X2 [cm]	Field Y [cm]	Y1 [cm]	Y2 [cm]	X [cm]	Y [cm]	Z [cm]	Calculated SSD [cm]	MU	Ref. D [Gy]
1	Field 1	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	181.0	10.0	0.0	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	91.3		

ANON96873


- Series
 - HCSWB V2.0
 - Etapa 1
 - IMRT
 - VMAT

- Open...
- Close
- Drop to view
- Export
- Plan Approval
- Paste Plan
- Copy Plan**
- Delete
- Delete Plan from Plan Sum
- Clear DVH Estimates
- Delete Trade-Off Plan Collection
- Properties


VMAT - Unapproved - Transversal - HCSWB V2.0




3D DVH BEV Arc VMAT - Unapproved - Model View - HCSWB V2.0



VMAT - Unapproved - Frontal - HCSWB V2.0



VMAT - Unapproved - Sagittal - HCSWB V2.0



Group	Field ID	Technique	Machine/Energy	MLC	Field Weight	Scale	Gantry Rtn [deg]	Coll Rtn [deg]	Couch Rtn [deg]	Wedge	Field X [cm]	X1 [cm]	X2 [cm]	Field Y [cm]	Y1 [cm]	Y2 [cm]	X [cm]	Y [cm]	Z [cm]	Calculated SSD [cm]	MU	Ref. D [Gy]
I	Field 1	ARC-I	Eclipse CAP - 6X		1.000	IEC61217	181.0 CW 179.0	10.0	0.0	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	91.3		
I	Field 2	ARC-I	Fclipse CAP - 6X		1.000	IEC61217	181.0 CW 179.0	350.0	0.0	None	18.2	-8.9	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	91.3		

ANON96873


- Series
 - HCSWB V2.0
 - Etapa 1
 - IMRT
 - VMAT

IMRT

- HCSWB V2.0
 - Registered Images
 - HCSWB V2.0
 - User Origin
 - Reference Points
 - PTV_WBopt
 - Dose
 - Fields
 - Isocenter Group I
 - Field 1

Structures and Objectives

Structure	Volume [cc]	Points	Resolution [mm]
<input checked="" type="checkbox"/> _Brain&BODY	959	31953	3.00
<input checked="" type="checkbox"/> _Brainstem#Hi	19	2000	2.06
<input checked="" type="checkbox"/> _Eyes&Body	916	30546	3.00



Calculation Options

Model PGO_15161: Plan Geometry Optimizer (version 15.1.61)
Plan geometry optimization algorithm for photon beams.

GEOS calculation options	Non-coplanar
Initial field distribution	
Initial number of fields	71
Maximum collimator variation	0
Coplanar offset angle	0
Maximum elevation angle for non coplanar fields	90
Minimum number of fields	9
Maximum number of fields	10
Fluence iterations per global geometric iteration	3
Local geometric optimization mode	Simplex
Fluence iterations per local geometric iteration	3
Maximum number of local optimization iterations	10
Initial step size in local optimization	3
Minimum field separation angle	10

Maximum number of fields

Min number of fields:

Max number of fields:

Continue to Optimization

Group	Field ID	Technique	Machine/Energy
I	Field 1	STATIC-I	Eclipse CAP - 6X

1.000	IEC61217	181.0	10.0	0.0	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24
-------	----------	-------	------	-----	------	------	------	------	------	------	------	------

Y [cm] Z [cm] Calculated SSD [cm] MU Ref. D [Gy]


1.58 1.09 91.3

Beam Angle Optimization - Cerebro, HCSWB_SIB (HCSWB)

Structures and Objectives

Structure	Volume [cc]	Points	Resolution [mm]
_Brain&BODY	959	180068	3.00
_Brainstem#Hi	19	10072	2.06
_Eyes&Body	916	90674	3.00
BODY	6255	242914	4.50
Brain	1242	103178	3.00
Brainstem	21	10295	2.12
Chiasm	0	2569	1.00
Cord	16	14927	1.93
Eyes	14	10456	1.84
Hippo+05	0.0	700.0	50
Hippocampus	31	12347	2.40
Leye	40.0	900.0	50
LLacrimal	4	14953	1.25
LLens	7	8624	1.45
Loptic	0.0	700.0	50
NS_Artifact	0	1960	1.00
PTV_WB	0	1189	1.00
PTV_WBopt	0.0	500.0	50
Reye	1	4514	1.00
RLacrimal	7	14280	1.47
RLens	1212	108099	3.00
Roptic			


Add Upper Objective Add Lower Objective Delete Objective



Irregular Surface Compensator PO_15606 Not supported for this plan

Stereotactic Dose

Dose Volume Histogram



Performing GLOBAL optimization... 61/71 fields left

Optimizing 0h 0m 15s 0

Calculation Options

Global optimization: Non-coplanar

Local optimization: Simplex

Min number of fields: 9

Max number of fields: 10

Optimize

Continue to Optimization

HCSWB

- Etapa 1
 - IMRT
 - VMAT

IMRT

HCSWB V2.0

Registered Images

- HCSWB V2.0
 - _Brain&BODY
 - _Brainstem#Hi
 - _Eyes&Body
 - BODY
 - Brain
 - Brainstem
 - Chiasm
 - Cord
 - Eyes
 - Hippo+05
 - Hippocampus
 - Leye
 - LLacrimal
 - LLens
 - Loptic
 - NS_Artifact
 - PTV_WB
 - PTV_WBopt
 - Reye
 - RLacrimal
 - RLens
 - Roptic



Beam Angle Optimization - Cerebro, HCSWB_SIB (HCSWB)

Structures and Objectives

Use Normal Tissue Objective Priority: 100 Define Settings...

Structure	Volume [cc]	Points	Resolution [mm]
<input checked="" type="checkbox"/> _Brain&BODY	959	180068	3.00
<input checked="" type="checkbox"/> _Brainstem#Hi	19	10072	2.06
<input checked="" type="checkbox"/> _Eyes&Body	916	90674	3.00
<input checked="" type="checkbox"/> BODY	6255	242914	4.50
<input checked="" type="checkbox"/> Brain	1242	103178	3.00
<input checked="" type="checkbox"/> Brainstem	21	10295	2.12
<input checked="" type="checkbox"/> Chiasm	0	2569	1.00
<input checked="" type="checkbox"/> Cord	16	14927	1.93
<input checked="" type="checkbox"/> Eyes	14	10456	1.84
Upper	Volume [%]: 0.0	Dose [Gy]: 700.0	Priority: 50
Hippo+05	Volume [cc]: 31	Points: 12347	Resolution [mm]: 2.40
Upper	Volume [%]: 0.0	Dose [Gy]: 1100.0	Priority: 50
Upper	Volume [%]: 40.0	Dose [Gy]: 900.0	Priority: 50
<input checked="" type="checkbox"/> Hippocampus	Volume [cc]: 4	Points: 14953	Resolution [mm]: 1.25
<input checked="" type="checkbox"/> Leye	Volume [cc]: 7	Points: 8624	Resolution [mm]: 1.45
Upper	Volume [%]: 0.0	Dose [Gy]: 700.0	Priority: 50
<input checked="" type="checkbox"/> LLacrimal	Volume [cc]: 0	Points: 1960	Resolution [mm]: 1.00
<input checked="" type="checkbox"/> LLens	Volume [cc]: 0	Points: 1189	Resolution [mm]: 1.00
Upper	Volume [%]: 0.0	Dose [Gy]: 500.0	Priority: 50
<input checked="" type="checkbox"/> Loptic	Volume [cc]: 1	Points: 4514	Resolution [mm]: 1.00
<input checked="" type="checkbox"/> NS_Artifact	Volume [cc]: 7	Points: 14280	Resolution [mm]: 1.47
<input checked="" type="checkbox"/> PTV_WB	Volume [cc]: 1212	Points: 108099	Resolution [mm]: 3.00

Add Upper Objective Add Lower Objective Delete Objective



Performing GLOBAL optimization...
35/71 fields left

Calculation Options

Global optimization: Non-coplanar

Local optimization: Simplex

Min number of fields:

Max number of fields:



Continue to Optimization

Beam Angle Optimization - Cerebro, HCSWB_SIB (HCSWB)

Structures and Objectives

Structure	Volume [cc]	Points	Resolution [mm]	Priority
<input checked="" type="checkbox"/> _Brain&BODY	959	180068	3.00	
<input checked="" type="checkbox"/> _Brainstem#Hi	19	10072	2.06	
<input checked="" type="checkbox"/> _Eyes&Body	916	90674	3.00	
<input checked="" type="checkbox"/> BODY	6255	242914	4.50	
<input checked="" type="checkbox"/> Brain	1242	103178	3.00	
<input checked="" type="checkbox"/> Brainstem	21	10295	2.12	
<input checked="" type="checkbox"/> Chiasm	0	2569	1.00	
<input checked="" type="checkbox"/> Cord	16	14927	1.93	
<input checked="" type="checkbox"/> Eyes	14	10456	1.84	
Upper	Volume [%]: 0.0	Dose [Gy]: 700.0	Priority: 50	
Hippo+05	Volume [cc]: 31	Points: 12347	Resolution [mm]: 2.40	
Upper	Volume [%]: 0.0	Dose [Gy]: 1100.0	Priority: 50	
Upper	Volume [%]: 40.0	Dose [Gy]: 900.0	Priority: 50	
<input checked="" type="checkbox"/> Hippocampus	Volume [cc]: 4	Points: 14953	Resolution [mm]: 1.25	
<input checked="" type="checkbox"/> Leye	Volume [cc]: 7	Points: 8624	Resolution [mm]: 1.45	
Upper	Volume [%]: 0.0	Dose [Gy]: 700.0	Priority: 50	
<input checked="" type="checkbox"/> LLacrimal	Volume [cc]: 0	Points: 1960	Resolution [mm]: 1.00	
<input checked="" type="checkbox"/> LLens	Volume [cc]: 0	Points: 1189	Resolution [mm]: 1.00	
Upper	Volume [%]: 0.0	Dose [Gy]: 500.0	Priority: 50	
<input checked="" type="checkbox"/> Loptic	Volume [cc]: 1	Points: 4514	Resolution [mm]: 1.00	
<input checked="" type="checkbox"/> NS_Artifact	Volume [cc]: 7	Points: 14280	Resolution [mm]: 1.47	
<input checked="" type="checkbox"/> PTV_WB	Volume [cc]: 1212	Points: 108099	Resolution [mm]: 3.00	

Add Upper Objective Add Lower Objective Delete Objective



Performing GLOBAL optimization...
10/71 fields left

Optimizing ... 0h 1m 43s 6

Calculation Options

Global optimization:
Non-coplanar

Local optimization:
Simplex

Min number of fields: 9
Max number of fields: 10

Edit...

OPTIMIZE

Continue to Optimization

OK Cancel Apply

HCSWB

- Etapa 1
 - IMRT
 - VMAT

IMRT

- HCSWB V2.0
 - Registered Images
 - HCSWB V2.0
 - _Brain&BODY
 - _Brainstem#Hi
 - _Eyes&Body
 - BODY
 - Brain
 - Brainstem
 - Chiasm
 - Cord
 - Eyes
 - Hippo+05
 - Hippocampus
 - Leye
 - LLacrimonal
 - LLens
 - Loptic
 - NS_Artifact
 - PTV_WB
 - PTV_WBopt
 - Reye
 - RLacrimonal
 - RLens
 - Roptic


Beam Angle Optimization - Cerebro, HCSWB_SIB (HCSWB)

Structures and Objectives


Use Normal Tissue Objective Priority: 100 Define Settings...

Structure	Volume [cc]	Points	Resolution [mm]
<input checked="" type="checkbox"/> _Brain&BODY	959	180068	3.00
<input checked="" type="checkbox"/> _Brainstem#Hi	19	10072	2.06
<input checked="" type="checkbox"/> _Eyes&Body	916	90674	3.00
<input checked="" type="checkbox"/> BODY	6255	242914	4.50
<input checked="" type="checkbox"/> Brain	1242	103178	3.00
<input checked="" type="checkbox"/> Brainstem	21	10295	2.12
<input checked="" type="checkbox"/> Chiasm	0	2569	1.00
<input checked="" type="checkbox"/> Cord	16	14927	1.93
<input checked="" type="checkbox"/> Eyes	14	10456	1.84
Upper	Volume [%]: 0.0	Dose [Gy]: 700.0	Priority: 50
Hippo+05	Volume [cc]: 31	Points: 12347	Resolution [mm]: 2.40
Upper	Volume [%]: 0.0	Dose [Gy]: 1100.0	Priority: 50
Upper	Volume [%]: 40.0	Dose [Gy]: 900.0	Priority: 50
<input checked="" type="checkbox"/> Hippocampus	Volume [cc]: 4	Points: 14953	Resolution [mm]: 1.25
<input checked="" type="checkbox"/> Leye	Volume [cc]: 7	Points: 8624	Resolution [mm]: 1.45
Upper	Volume [%]: 0.0	Dose [Gy]: 700.0	Priority: 50
<input checked="" type="checkbox"/> LLacrimonal	Volume [cc]: 0	Points: 1960	Resolution [mm]: 1.00
Upper	Volume [%]: 0.0	Dose [Gy]: 500.0	Priority: 50
<input checked="" type="checkbox"/> Loptic	Volume [cc]: 1	Points: 4514	Resolution [mm]: 1.00
<input checked="" type="checkbox"/> NS_Artifact	Volume [cc]: 7	Points: 14280	Resolution [mm]: 1.47
<input checked="" type="checkbox"/> PTV_WB	Volume [cc]: 1212	Points: 108099	Resolution [mm]: 3.00

Add Upper Objective Add Lower Objective Delete Objective



Irregular Surface Compensator PO_15606 Not supported for this plan
Stereotactic Dose



Calculation Options

0 iterations left

Global optimization: Non-coplanar

Local optimization: Simplex

Min number of fields: 9

Max number of fields: 10

Continue Optimization


OK Cancel Apply

HCSWB

- Etapa 1
 - IMRT
 - VMAT


IMRT

- HCSWB V2.0
 - Registered Images
 - HCSWB V2.0
 - User Origin
 - Reference Points
 - PTV_WBopt
 - Dose
 - Fields
 - Isocenter Group I
 - Field 1
 - Field 2
 - Field 3
 - Field 4
 - Field 5
 - Field 6
 - Field 7
 - Field 8
 - Field 9



Beam	Couch angle (°)	Gantry angle (°)
1	140	150
2	150	230
3	225	360
4	190	76
5	196	131
6	96	171
7	150	275
8	196	223
9	90	221






Haz	Varian		IEC	
	Camilla	Gantry	Camilla	Gantry
1	140	150	40	30
2	150	230	30	310
3	255	0	285	180
4	190	76	350	104
5	196	131	344	49
6	96	171	84	9
7	150	275	30	265
8	196	223	344	317
9	90	221	90	319



Group	Field ID	Technique	Machine/Energy	MLC	Field Weight	Scale	Gantry Rtn [deg]	Coll Rtn [deg]	Couch Rtn [deg]	Wedge	Field X [cm]	X1 [cm]	X2 [cm]	Field Y [cm]	Y1 [cm]	Y2 [cm]	X [cm]	Y [cm]	Z [cm]	Calculated SSD [cm]	MU	Ref. D [Gy]
I	Field 1	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	80.3	0.0	11.9	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	92.1		
I	Field 2	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	297.5	0.0	333.7	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	90.4		
I	Field 3	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	227.4	0.0	324.0	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	91.3		
I	Field 4	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	142.8	0.0	72.1	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	90.6		
I	Field 5	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	272.1	0.0	4.2	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	92.6		
I	Field 6	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	320.0	0.0	303.0	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	87.1		
I	Field 7	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	310.0	0.0	270.8	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	83.8		
I	Field 8	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	202.4	0.0	36.8	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	91.7		
I	Field 9	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	298.2	0.0	306.0	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	87.3		

Varian Coordinate System

IEC Coordinate System



X1, X2, : \longleftrightarrow
Y1, Y2 : -1 0 +1

X1, Y1: 1 0 -1
X2, Y2: -1 0 1

Beam

Couch angle (°)

Gantry angle (°)

1	140	150
2	150	230
3	225	360
4	190	76
5	196	131
6	96	171
7	150	275
8	196	223
9	90	221




Haz	Varian		IEC	
	Camilla	Gantry	Camilla	Gantry
1	140	150	40	30
2	150	230	30	310
3	255	0	285	180
4	190	76	350	104
5	196	131	344	49
6	96	171	84	9
7	150	275	30	265
8	196	223	344	317
9	90	221	90	319

HCSWB

- Etapa 1
 - IMRT
 - VMAT

IMRT

- HCSWB V2.0
 - Registered Images
 - HCSWB V2.0
 - User Origin
 - Reference Points
 - PTV_WBopt
 - Dose
 - Fields
 - Isocenter Group 1
 - Field 1
 - Field 2
 - Field 3
 - Field 4
 - Field 5
 - Field 6
 - Field 7
 - Field 8
 - Field 9



Group	Field ID	Technique	Machine/Energy	MLC	Field Weight	Scale	Gantry Rtn [deg]	Coll Rtn [deg]	Couch Rtn [deg]	Wedge	Field X [cm]	X1 [cm]	X2 [cm]	Field Y [cm]	Y1 [cm]	Y2 [cm]	X [cm]	Y [cm]	Z [cm]	Calculated SSD [cm]	MU	Ref. D [Gy]
1	Field 1	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	80.3	0.0	11.9	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	92.1		
1	Field 2	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	297.5	0.0	333.7	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	90.4		
1	Field 3	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	227.4	0.0	324.0	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	91.3		
1	Field 4	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	142.8	0.0	72.1	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	90.6		
1	Field 5	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	272.1	0.0	4.2	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	92.6		
1	Field 6	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	320.0	0.0	303.0	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	87.1		
1	Field 7	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	310.0	0.0	270.8	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	83.8		
1	Field 8	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	202.4	0.0	36.8	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	91.7		
1	Field 9	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	298.2	0.0	306.0	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	87.3		

QuickLinks Cerebro, HCSWB_SIB (HCSWB) Worklist

File Edit View Insert Planning Tools Window




2.0 cm 2.0 cm 1

HCSWB

- Etapa 1
 - IMRT
 - VMAT

IMRT

- HCSWB V2.0
 - Registered Images
 - HCSWB V2.0
 - User Origin
 - Reference Points
 - PTV_WBopt
 - Dose
 - Fields
 - Isocenter Group 1
 - Field 1
 - Field 2
 - Field 3
 - Field 4
 - Field 5
 - Field 6
 - Field 7
 - Field 8
 - Field 9






Group	Field ID	Technique	Machine/Energy	MLC	Field Weight	Scale	Gantry Rtn [deg]	Coll Rtn [deg]	Couch Rtn [deg]	Wedge	Field X [cm]	X1 [cm]	X2 [cm]	Field Y [cm]	Y1 [cm]	Y2 [cm]	X [cm]	Y [cm]	Z [cm]	Calculated SSD [cm]	MU	Ref. D [Gy]
1	Field 1	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	80.3	0.0	11.9	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	92.1		
1	Field 2	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	297.5	0.0	333.7	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	90.4		
1	Field 3	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	227.4	0.0	324.0	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	91.3		
1	Field 4	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	142.8	0.0	72.1	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	90.6		
1	Field 5	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	272.1	0.0	4.2	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	92.6		
1	Field 6	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	320.0	0.0	303.0	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	87.1		
1	Field 7	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	310.0	0.0	270.8	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	83.8		
1	Field 8	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	202.4	0.0	36.8	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	91.7		
1	Field 9	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	298.2	0.0	306.0	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	87.3		

HCSWB

- Etapa 1
 - IMRT
 - VMAT

IMRT

- HCSWB V2.0
 - Registered Images
 - HCSWB V2.0
 - User Origin
 - Reference Points
 - PTV_WBopt
 - Dose
 - Fields
 - Isocenter Group 1
 - Field 1
 - Field 2
 - Field 3
 - Field 4
 - Field 5
 - Field 6
 - Field 7
 - Field 8
 - Field 9







Group	Field ID	Technique	Machine/Energy	MLC	Field Weight	Scale	Gantry Rtn [deg]	Coll Rtn [deg]	Couch Rtn [deg]	Wedge	Field X [cm]	X1 [cm]	X2 [cm]	Field Y [cm]	Y1 [cm]	Y2 [cm]	X [cm]	Y [cm]	Z [cm]	Calculated SSD [cm]	MU	Ref. D [Gy]
1	Field 1	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	80.3	0.0	11.9	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	92.1		
1	Field 2	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	297.5	0.0	333.7	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	90.4		
1	Field 3	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	227.4	0.0	324.0	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	91.3		
1	Field 4	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	142.8	0.0	72.1	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	90.6		
1	Field 5	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	272.1	0.0	4.2	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	92.6		
1	Field 6	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	320.0	0.0	303.0	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	87.1		
1	Field 7	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	310.0	0.0	270.8	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	83.8		
1	Field 8	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	202.4	0.0	36.8	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	91.7		
1	Field 9	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	298.2	0.0	306.0	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	87.3		

HCSWB

- Etapa 1
 - IMRT
 - VMAT

IMRT

- HCSWB V2.0
 - Registered Images
 - HCSWB V2.0
 - User Origin
 - Reference Points
 - PTV_WBopt
 - Dose
 - Fields
 - Isocenter Group 1
 - Field 1
 - Field 2
 - Field 3
 - Field 4
 - Field 5
 - Field 6
 - Field 7
 - Field 8
 - Field 9







Group	Field ID	Technique	Machine/Energy	MLC	Field Weight	Scale	Gantry Rtn [deg]	Coll Rtn [deg]	Couch Rtn [deg]	Wedge	Field X [cm]	X1 [cm]	X2 [cm]	Field Y [cm]	Y1 [cm]	Y2 [cm]	X [cm]	Y [cm]	Z [cm]	Calculated SSD [cm]	MU	Ref. D [Gy]
1	Field 1	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	80.3	0.0	11.9	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	92.1		
1	Field 2	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	297.5	0.0	333.7	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	90.4		
1	Field 3	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	227.4	0.0	324.0	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	91.3		
1	Field 4	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	142.8	0.0	72.1	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	90.6		
1	Field 5	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	272.1	0.0	4.2	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	92.6		
1	Field 6	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	320.0	0.0	303.0	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	87.1		
1	Field 7	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	310.0	0.0	270.8	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	83.8		
1	Field 8	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	202.4	0.0	36.8	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	91.7		
1	Field 9	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	298.2	0.0	306.0	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	87.3		

HCSWB

- Etapa 1
 - IMRT
 - VMAT

IMRT

- HCSWB V2.0
 - Registered Images
 - HCSWB V2.0
 - User Origin
 - Reference Points
 - PTV_WBopt
 - Dose
 - Fields
 - Isocenter Group 1
 - Field 1
 - Field 2
 - Field 3
 - Field 4
 - Field 5
 - Field 6
 - Field 7
 - Field 8
 - Field 9



Group	Field ID	Technique	Machine/Energy	MLC	Field Weight	Scale	Gantry Rtn [deg]	Coll Rtn [deg]	Couch Rtn [deg]	Wedge	Field X [cm]	X1 [cm]	X2 [cm]	Field Y [cm]	Y1 [cm]	Y2 [cm]	X [cm]	Y [cm]	Z [cm]	Calculated SSD [cm]	MU	Ref. D [Gy]
I	Field 1	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	80.3	0.0	11.9	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	92.1		
I	Field 2	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	297.5	0.0	333.7	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	90.4		
I	Field 3	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	227.4	0.0	324.0	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	91.3		
I	Field 4	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	142.8	0.0	72.1	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	90.6		
I	Field 5	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	272.1	0.0	4.2	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	92.6		
I	Field 6	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	320.0	0.0	303.0	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	87.1		
I	Field 7	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	310.0	0.0	270.8	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	83.8		
I	Field 8	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	202.4	0.0	36.8	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	91.7		
I	Field 9	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	298.2	0.0	306.0	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	87.3		

QuickLinks Cerebro, HCSWB_SIB (HCSWB) Worklist

File Edit View Insert Planning Tools Window





2.0 cm 2.0 cm 1

HCSWB

- Etapa 1
 - IMRT
 - VMAT

IMRT

- HCSWB V2.0
 - Registered Images
 - HCSWB V2.0
 - User Origin
 - Reference Points
 - PTV_WBopt
 - Dose
 - Fields
 - Isocenter Group 1
 - Field 1
 - Field 2
 - Field 3
 - Field 4
 - Field 5
 - Field 6
 - Field 7
 - Field 8
 - Field 9







Group	Field ID	Technique	Machine/Energy	MLC	Field Weight	Scale	Gantry Rtn [deg]	Coll Rtn [deg]	Couch Rtn [deg]	Wedge	Field X [cm]	X1 [cm]	X2 [cm]	Field Y [cm]	Y1 [cm]	Y2 [cm]	X [cm]	Y [cm]	Z [cm]	Calculated SSD [cm]	MU	Ref. D [Gy]
I	Field 1	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	80.3	0.0	11.9	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	92.1		
I	Field 2	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	297.5	0.0	333.7	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	90.4		
I	Field 3	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	227.4	0.0	324.0	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	91.3		
I	Field 4	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	142.8	0.0	72.1	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	90.6		
I	Field 5	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	272.1	0.0	4.2	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	92.6		
I	Field 6	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	320.0	0.0	303.0	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	87.1		
I	Field 7	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	310.0	0.0	270.8	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	83.8		
I	Field 8	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	202.4	0.0	36.8	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	91.7		
I	Field 9	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	298.2	0.0	306.0	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	87.3		

HCSWB

- Etapa 1
 - IMRT
 - VMAT

IMRT

- HCSWB V2.0
 - Registered Images
 - HCSWB V2.0
 - User Origin
 - Reference Points
 - PTV_WBopt
 - Dose
 - Fields
 - Isocenter Group 1
 - Field 1
 - Field 2
 - Field 3
 - Field 4
 - Field 5
 - Field 6
 - Field 7
 - Field 8
 - Field 9



Fields Dose Field Alignments Plan Objectives Optimization Objectives Dose Statistics Reference Points Calculation Models Plan Sum

Group	Field ID	Technique	Machine/Energy	MLC	Field Weight	Scale	Gantry Rtn [deg]	Coll Rtn [deg]	Couch Rtn [deg]	Wedge	Field X [cm]	X1 [cm]	X2 [cm]	Field Y [cm]	Y1 [cm]	Y2 [cm]	X [cm]	Y [cm]	Z [cm]	Calculated SSD [cm]	MU	Ref. D [Gy]
1	Field 1	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	80.3	0.0	11.9	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	92.1		
1	Field 2	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	297.5	0.0	333.7	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	90.4		
1	Field 3	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	227.4	0.0	324.0	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	91.3		
1	Field 4	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	142.8	0.0	72.1	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	90.6		
1	Field 5	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	272.1	0.0	4.2	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	92.6		
1	Field 6	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	320.0	0.0	303.0	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	87.1		
1	Field 7	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	310.0	0.0	270.8	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	83.8		
1	Field 8	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	202.4	0.0	36.8	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	91.7		
1	Field 9	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	298.2	0.0	306.0	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	87.3		

QuickLinks Cerebro, HCSWB_SIB (HCSWB) Worklist

File Edit View Insert Planning Tools Window





2.0 cm 2.0 cm 1

HCSWB

- Etapa 1
 - IMRT
 - VMAT

IMRT

- HCSWB V2.0
 - Registered Images
 - HCSWB V2.0
 - User Origin
 - Reference Points
 - PTV_WBopt
 - Dose
 - Fields
 - Isocenter Group 1
 - Field 1
 - Field 2
 - Field 3
 - Field 4
 - Field 5
 - Field 6
 - Field 7
 - Field 8
 - Field 9





Group	Field ID	Technique	Machine/Energy	MLC	Field Weight	Scale	Gantry Rtn [deg]	Coll Rtn [deg]	Couch Rtn [deg]	Wedge	Field X [cm]	X1 [cm]	X2 [cm]	Field Y [cm]	Y1 [cm]	Y2 [cm]	X [cm]	Y [cm]	Z [cm]	Calculated SSD [cm]	MU	Ref. D [Gy]
I	Field 1	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	80.3	0.0	11.9	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	92.1		
I	Field 2	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	297.5	0.0	333.7	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	90.4		
I	Field 3	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	227.4	0.0	324.0	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	91.3		
I	Field 4	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	142.8	0.0	72.1	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	90.6		
I	Field 5	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	272.1	0.0	4.2	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	92.6		
I	Field 6	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	320.0	0.0	303.0	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	87.1		
I	Field 7	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	310.0	0.0	270.8	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	83.8		
I	Field 8	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	202.4	0.0	36.8	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	91.7		
I	Field 9	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	298.2	0.0	306.0	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	87.3		

HCSWB


- Etapa 1
 - IMRT
 - VMAT

IMRT

- HCSWB V2.0
 - Registered Images
 - HCSWB V2.0
 - User Origin
 - Reference Points
 - PTV_WBopt
 - Dose
 - Fields
 - Isocenter Group 1
 - Field 1
 - Field 2
 - Field 3
 - Field 4
 - Field 5
 - Field 6
 - Field 7
 - Field 8
 - Field 9



Group	Field ID	Technique	Machine/Energy	MLC	Field Weight	Scale	Gantry Rtn [deg]	Coll Rtn [deg]	Couch Rtn [deg]	Wedge	Field X [cm]	X1 [cm]	X2 [cm]	Field Y [cm]	Y1 [cm]	Y2 [cm]	X [cm]	Y [cm]	Z [cm]	Calculated SSD [cm]	MU	Ref. D [Gy]
1	Field 1	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	80.3	0.0	11.9	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	92.1		
1	Field 2	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	297.5	0.0	333.7	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	90.4		
1	Field 3	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	227.4	0.0	324.0	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	91.3		
1	Field 4	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	142.8	0.0	72.1	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	90.6		
1	Field 5	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	272.1	0.0	4.2	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	92.6		
1	Field 6	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	320.0	0.0	303.0	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	87.1		
1	Field 7	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	310.0	0.0	270.8	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	83.8		
1	Field 8	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	202.4	0.0	36.8	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	91.7		
1	Field 9	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	298.2	0.0	306.0	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	87.3		




Group	Field ID	Technique	Machine/Energy	MLC	Field Weight	Scale	Gantry Rtn [deg]	Coll Rtn [deg]	Couch Rtn [deg]	Wedge	Field X [cm]	X1 [cm]	X2 [cm]	Field Y [cm]	Y1 [cm]	Y2 [cm]	X [cm]	Y [cm]	Z [cm]	Calculated SSD [cm]	MU	Ref. D [Gy]
I	Field 1	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	80.3	0.0	11.9	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	92.1		
I	Field 2	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	297.5	0.0	333.7	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	90.4		
I	Field 3	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	227.4	0.0	324.0	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	91.3		
I	Field 4	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	142.8	0.0	72.1	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	90.6		
I	Field 5	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	272.1	0.0	4.2	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	92.6		
I	Field 6	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	320.0	0.0	303.0	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	87.1		
I	Field 7	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	310.0	0.0	270.8	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	83.8		
I	Field 8	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	202.4	0.0	36.8	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	91.7		
I	Field 9	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	298.2	0.0	306.0	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	87.3		

Opens the Optimization dialog.

User: Varianinstaller Group: System Administrator Site: Main CAP NUM SCRL

11:25 AM 10/26/2023

Plan Information						
ID/Type	cm ³	Vol [%]	Dose[Gy]	Actual Dose[Gy]	Priority	gEUD a
<input checked="" type="checkbox"/> PTV_WBopt	1192.7					
Upper	11.9	1.0	3060.00		100	x
Lower	1180.8	99.0	2940.00		100	x
<input checked="" type="checkbox"/> Eyes	13.8					
Upper	0.0	0.0	700.00		50	x
<input checked="" type="checkbox"/> Hippocampus	4.4					
Upper	0.0	0.0	1100.00		50	x
Upper	1.7	40.0	900.00		50	x
<input checked="" type="checkbox"/> Leye	6.8					
Upper	0.0	0.0	700.00		50	x
<input checked="" type="checkbox"/> LLens	0.1					
Upper	0.0	0.0	500.00		50	x
<input checked="" type="checkbox"/> Reye	7.0					
Upper	0.0	0.0	700.00		50	x
<input checked="" type="checkbox"/> RLens	0.2					
Upper	0.0	0.0	500.00		50	x
<input checked="" type="checkbox"/> _Brain&BODY	958.6					
<input checked="" type="checkbox"/> _Brainstem#H	19.5					
<input checked="" type="checkbox"/> _Eyes&Body	916.4					
<input checked="" type="checkbox"/> BODY	6255.0					
Normal Tissue Objective					Off	
Base Dose Plan					None	
Settings					1000/6000s/Normal (2.5 mm)	



Progress | Plan Objectives

- 3D Dose Max
- 3D MAX for PTVWBopt
- 3D MEAN for PTVWBopt
- 3D MIN for PTVWBopt
- Elapsed Time
- Iteration

- Fluence
- Intermediate Dose

Open Log...

Automatic Optimization Mode
 Automatic Intermediate Dose
 Start IMRT Optimization
 Intermediate Dose
 OK
 Cancel

Plan Information


ID/Type	cm ³	Vol [%]	Dose[Gy]	Actual Dose[Gy]	Priority	gEUD a
<input checked="" type="checkbox"/> PTV_WBopt	1192.7					
Upper	11.9	1.0	3060.00		100	x
Lower	1180.8	99.0	2940.00		100	x
<input checked="" type="checkbox"/> Eyes	13.8					
Upper	0.0	0.0	700.00		50	x
<input checked="" type="checkbox"/> Hippocampus	4.4					
Upper	0.0	0.0	1100.00		50	x
Upper	1.7	40.0	900.00		50	x
<input checked="" type="checkbox"/> Leye	6.8					
Upper	0.0	0.0	700.00		50	x
<input checked="" type="checkbox"/> LLens	0.1					
Upper	0.0	0.0	500.00		50	x

Normal Tissue Objective

Off
 Manual
 Automatic NTO
 Auto (SRS NTO)

Priority: 100
 Distance from Target Border: 0.50 cm
 Start Dose: 98.0 %
 End Dose: 30.0 %
 Fall-off: 0.10

Base Dose Plan: None
 Settings: 1000/6000s/Normal (2.5 mm)



Progress | Plan Objectives



- 3D Dose Max
- 3D MAX for PTVWBopt
- 3D MEAN for PTVWBopt
- 3D MIN for PTVWBopt
- Elapsed Time
- Iteration

Fluence
 Intermediate Dose

Open Log...

Automatic Optimization Mode
 Automatic Intermediate Dose

Plan Information							
ID/Type	cm ³	Vol [%]	Dose[Gy]	Actual Dose[Gy]	Priority	gEUD a	
<input checked="" type="checkbox"/> PTV_WBopt	1192.7						
Upper	11.9	1.0	3060.00	3187.51	100		x
Lower	1180.8	99.0	2940.00	2599.94	100		x
<input checked="" type="checkbox"/> Eyes	13.8						
Upper	0.0	0.0	700.00	1191.79	50		x
<input checked="" type="checkbox"/> Hippocampus	4.4						
Upper	0.0	0.0	1100.00	2154.10	50		x
Upper	1.7	40.0	900.00	1073.94	50		x
<input checked="" type="checkbox"/> Leye	6.8						
Upper	0.0	0.0	700.00	1439.29	50		x
<input checked="" type="checkbox"/> LLens	0.1						
Upper	0.0	0.0	500.00	490.17	50		x
<input checked="" type="checkbox"/> Reye	7.0						
Upper	0.0	0.0	700.00	1070.02	50		x
<input checked="" type="checkbox"/> RLens	0.2						
Upper	0.0	0.0	500.00	538.53	50		x
<input checked="" type="checkbox"/> _Brain&BODY	958.6						
<input checked="" type="checkbox"/> _Brainstem#H	19.5						
<input checked="" type="checkbox"/> _Eyes&Body	916.4						
<input checked="" type="checkbox"/> BODY	6255.0						
Normal Tissue Objective					100/Manual		
Base Dose Plan					None		
Settings					1000/6000s/Normal (2.5 mm)		





Progress	Plan Objectives
3D Dose Max	3592.40 Gy
3D MAX for PTVWBopt	3589.43 Gy
3D MEAN for PTVWBopt	2990.91 Gy
3D MIN for PTVWBopt	1396.48 Gy
Elapsed Time	8 s
Iteration	21
Fluence	
Intermediate Dose	

PTV_WBopt	<div style="width: 100%; height: 10px; background-color: red;"></div>
BODY	<div style="width: 100%; height: 10px; background-color: green;"></div>
Hippocampus	<div style="width: 100%; height: 10px; background-color: green;"></div>
Leye	<div style="width: 100%; height: 10px; background-color: cyan;"></div>
Reye	<div style="width: 100%; height: 10px; background-color: cyan;"></div>
Eyes	<div style="width: 100%; height: 10px; background-color: cyan;"></div>
Roptic	<div style="width: 100%; height: 10px; background-color: cyan;"></div>
RLens	<div style="width: 100%; height: 10px; background-color: cyan;"></div>

[Open Log...](#)

Automatic Optimization Mode
 Stop IMRT Optimization
 Intermediate Dose
 Optimizing...


Plan Information						
ID/Type	cm ³	Vol [%]	Dose[Gy]	Actual Dose[Gy]	Priority	gEUD a
<input checked="" type="checkbox"/> PTV_WBopt	1192.7					
Upper	0.0	0.0	3060.00	3263.39	100	x
Lower	1180.8	99.0	2940.00	2646.40	100	x
<input checked="" type="checkbox"/> Eyes	13.8					
Upper	0.0	0.0	700.00	1144.69	50	x
<input checked="" type="checkbox"/> Hippocampus	4.4					
Upper	0.0	0.0	1100.00	1997.60	50	x
Upper	1.7	40.0	900.00	1024.29	50	x
<input checked="" type="checkbox"/> Leye	6.8					
Upper	0.0	0.0	700.00	1657.48	50	x
<input checked="" type="checkbox"/> LLens	0.1					
Upper	0.0	0.0	500.00	466.03	50	x
<input checked="" type="checkbox"/> Reye	7.0					
Upper	0.0	0.0	700.00	976.42	50	x
<input checked="" type="checkbox"/> RLens	0.2					
Upper	0.0	0.0	500.00	528.38	50	x
<input checked="" type="checkbox"/> _Brain&BODY	958.6					
<input checked="" type="checkbox"/> _Brainstem#H	19.5					
<input checked="" type="checkbox"/> _Eyes&Body	916.4					
<input checked="" type="checkbox"/> BODY	6255.0					
Normal Tissue Objective					100/Manual	
Base Dose Plan					None	
Settings					1000/6000s/Normal (2.5 mm)	



Progress Plan Objectives

3D Dose Max 3268.57 Gy
 3D MAX for PTVWBopt 3263.39 Gy
 3D MEAN for PTVWBopt 2991.61 Gy
 3D MIN for PTVWBopt 1269.63 Gy
 Elapsed Time 83 s
 Iteration 224



Fluence
 Intermediate Dose



Automatic Optimization Mode
 Stop IMRT Optimization

 Optimization paused.


Plan Information						
ID/Type	cm ³	Vol [%]	Dose[Gy]	Actual Dose[Gy]	Priority	gEUD a
<input checked="" type="checkbox"/> PTV_WBopt	1192.7					
Upper	0.0	0.0	3060.00	3263.39	100	x
Lower	1180.8	99.0	2940.00	2646.40	100	x
<input checked="" type="checkbox"/> Eyes	13.8					
Upper	0.0	0.0	700.00	1144.69	50	x
<input checked="" type="checkbox"/> Hippocampus	4.4					
Upper	0.0	0.0	1100.00	1997.60	50	x
Upper	1.7	40.0	900.00	1024.29	50	x
<input checked="" type="checkbox"/> Leye	6.8					
Upper	0.0	0.0	700.00	1657.48	50	x
<input checked="" type="checkbox"/> LLens	0.1					
Upper	0.0	0.0	500.00	466.03	50	x
<input checked="" type="checkbox"/> Reye	7.0					
Upper	0.0	0.0	700.00	976.42	50	x
<input checked="" type="checkbox"/> RLens	0.2					
Upper	0.0	0.0	500.00	528.38	50	x
<input checked="" type="checkbox"/> _Brain&BODY	958.6					
<input checked="" type="checkbox"/> _Brainstem#H	19.5					
<input checked="" type="checkbox"/> _Eyes&Body	916.4					
<input checked="" type="checkbox"/> BODY	6255.0					
Normal Tissue Objective					100/Manual	
Base Dose Plan					None	
Settings					1000/6000s/Normal (2.5 mm)	



Progress | Plan Objectives

3D Dose Max: 3268.57 Gy
 3D MAX for PTVWBopt: 3263.39 Gy
 3D MEAN for PTVWBopt: 2991.61 Gy
 3D MIN for PTVWBopt: 1269.63 Gy
 Elapsed Time: 83 s
 Iteration: 224

Fluence
 Intermediate Dose



Automatic Optimization Mode
 Stop IMRT Optimization

 Optimization paused.

- Arc Geometry Tool... Ctrl+F7
- HyperArc
- Optimization
- Dose Calculation
 - Calculate Volume F5
 - Calculate Volume with Preset Values Shift+F5
 - Calculate Plane F6
 - Calculate Portal Dose Shift+F6
 - Calculate Leaf Motions... F11
 - Show/Edit Calculation Volume
 - Reset Calculation Volume
- Create Verification Plan...
- Create Partial Treatment Plan...
- Plan Normalization...
- Isodose Levels...
- Compensator Isolevels...
- Plan Uncertainty Parameters...
- Show Dose Volume Histogram View
- Create Plan Comparison DVH...
- Biological Evaluation...
- Verify MLC Leaf Positions...
- Field Weight... F3
- Change Treatment Units...
- DVH Based Plan Converter...
- Merge Subfields
- Edit Compensator...
- Field Order...
- Reference Point Organizer...
- Delta Couch Shift Editor...
- Templates and Clinical Protocols
- Enable Digitizer
- Check IHE-RO RT Compliance

Group	Field ID	Technique	Machine/Energy	MLC	Field Weight	Scale	Gantry Rtn [deg]	Coll Rtn [deg]	Couch Rtn [deg]	Wedge	Field X [cm]	X1 [cm]	X2 [cm]	Field Y [cm]	Y1 [cm]	Y2 [cm]	X [cm]	Y [cm]	Z [cm]	Calculated SSD [cm]	MU	Ref. D [Gy]
I	Field 1	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	80.3	0.0	11.9	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	92.1		
I	Field 2	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	297.5	0.0	333.7	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	90.4		
I	Field 3	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	227.4	0.0	324.0	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	91.3		
I	Field 4	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	142.8	0.0	72.1	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	90.6		
I	Field 5	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	272.1	0.0	4.2	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	92.6		
I	Field 6	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	320.0	0.0	303.0	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	87.1		
I	Field 7	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	310.0	0.0	270.8	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	83.8		
I	Field 8	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	202.4	0.0	36.8	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	91.7		
I	Field 9	STATIC-I	Eclipse CAP - 6X		1.000	IEC61217	298.2	0.0	306.0	None	18.0	-8.7	+9.3	14.7	-8.0	+6.7	0.24	1.58	1.09	87.3		

QuickLinks Cerebro, HCSWB_SIB (HCSWB) Worklist

File Edit View Insert Planning Tools Window




2.0 cm 2.0 cm 1

HCSWB

- Etapa 1
 - IMRT
 - VMAT

HCSWB V2.0

- User Origin
- Reference Points
 - PTV_WBopt
- Dose
- Fields
 - Isocenter Group I
 - Field 1
 - Field 2
 - Field 3
 - Field 4
 - Field 5
 - Field 6
 - Field 7
 - Field 8
 - Field 9



Fields Dose Field Alignments Plan Objectives Optimization Objectives Dose Statistics Reference Points Calculation Models Plan Sum

Use Default Models	Particle Type	Calculation Type	Calculation Model	Status	Algorithm	Calculation Options	
Clear All Selections	Photon	Volume Dose	AcurosXB_15606	OK	Acuros External Beam (Version 15.6.06)	Edit	
		DVH Estimation	DVH Estimation Algorithm [15.6.06]	OK	DVH Estimation Algorithm (Version 15.6.06)	Edit	
		Beam Angle Optimization	PGO_15161	OK	Plan Geometry Optimizer (Version 15.1.61)	Edit	
		IMRT Optimization	PO_15606	OK	Photon Optimizer (Version 15.6.06)	Edit	
		VMAT Optimization	PO_15606	OK			
		Irregular Surface Compensator	PO_15606	Not supported for this plan			
		Stereotactic Dose					Edit
		Portal Dose					

HCSWB

- Etapa 1
 - IMRT
 - VMAT

HCSWB V2.0

- User Origin
- Reference Points
 - PTV_WBopt
- Dose
- Fields
 - Isocenter Group 1
 - Field 1
 - Field 2
 - Field 3
 - Field 4
 - Field 5
 - Field 6
 - Field 7
 - Field 8
 - Field 9

IMRT - Unapproved - Transversal - HCSWB V2.0

3D DVH BEV Arc IMRT - Unapproved - BEV - SAD 100 cm - Field 9 - HCSWB V2.0


Calculation Options

Model AcurosXB_15606: Acuros External Beam (version 15.6.06)
AcurosXB: dose calculation algorithm for photon beams.

Acuros calculation options	Value
Calculation resolution in cm	0.25
Calculation resolution in cm for SRS and HyperArc™	0.125
Field normalization type	100% to isocenter
Dose reporting mode	Dose to medium
Heterogeneity correction	ON
Plan dose calculation	ON
Use GPU	Yes
Automatic high-density material	Bone
Maximum automatic high-density volume in cm ³	0.5

Use GPU
Defines whether to use the graphics processing unit (GPU) to accelerate the algorithm. Calculation hosts need to have specific hardware to execute GPU version. Also, there may be cases when calculation will require more memory than is available on the GPU card.

OK Cancel



Use Default Models	Particle Type	Calculation type	Calculation model	Status	Algorithm	Calculation Options
Clear All Selections	Photon	Volume Dose	AcurosXB_15606	OK	Acuros External Beam (Version 15.6.06)	Edit
		DVH Estimation	DVH Estimation Algorithm [15.6.06]	OK	DVH Estimation Algorithm (Version 15.6.06)	Edit
		Beam Angle Optimization	PGO_15161	OK	Plan Geometry Optimizer (Version 15.1.61)	Edit
		IMRT Optimization	PO_15606	OK	Photon Optimizer (Version 15.6.06)	Edit
		VMAT Optimization	PO_15606	OK		
		Irregular Surface Compensator	PO_15606	Not supported for this plan		
		Stereotactic Dose				
		Portal Dose				Edit

HCSWB

- Etapa 1
 - IMRT
 - VMAT

HCSWB V2.0

- User Origin
- Reference Points
 - PTV_WBopt
- Dose
- Fields
 - Isocenter Group 1
 - Field 1
 - Fluence
 - Field 2
 - Fluence
 - Field 3
 - Fluence
 - Field 4
 - Fluence
 - Field 5
 - Fluence
 - Field 6
 - Fluence
 - Field 7
 - Fluence
 - Field 8
 - Fluence
 - Field 9
 - Fluence

IMRT - Unapproved - Transversal - HCSWB V2.0

IMRT - Unapproved - BEV - SAD 100 cm - Field 9 - HCSWB V2.0

Calculation Options

Model Varian Leaf Motion Calculator [15.6.06]: Varian Leaf Motion Calculator (version 15.6.06)
Leaf Motion Calculator for Sliding Window and Multiple Static Segments delivery techniques.

LMCV calculation options

- Sliding window options
 - Normal mode options
 - Fixed mode options
- Multiple static segment options

Default delivery method

Sliding Window (SW)

Sliding Window (SW)

Multiple Static Segments (MSS)

None

Default delivery method

Default method to use for IMRT delivery. This option only affects the default setting. Each field in the plan can still define the method separately.

OK Cancel





Use Default Models	Particle Type	Calculation type	Calculation model	Status	Algorithm	Calculation Options	
Clear All Selections	Photon	Volume Dose	AcurosXB_15606	OK	Acuros External Beam (Version 15.6.06)	Edit	
		DVH Estimation	DVH Estimation Algorithm [15.6.06]	OK	DVH Estimation Algorithm (Version 15.6.06)	Edit	
		Beam Angle Optimization	PGO_15161	OK	Plan Geometry Optimizer (Version 15.1.61)	Edit	
		IMRT Optimization	PO_15606	OK	Photon Optimizer (Version 15.6.06)	Edit	
		VMAT Optimization	PO_15606	OK			
		Irregular Surface Compensator	PO_15606	Not supported for this plan			
		Stereotactic Dose					Edit
		Portal Dose					

HCSWB

- Etapa1
 - IMRT
 - VMAT

IMRT

- HCSWB V2.0
 - Registered Images
 - HCSWB V2.0
 - _Brain&BODY
 - _Brainstem#Hi
 - _Eyes&Body
 - BODY
 - Brain
 - Brainstem
 - Chiasm
 - Cord
 - Eyes
 - Hippo+05
 - Hippocampus
 - Leye
 - LLacrimal
 - LLens
 - Loptic
 - NS_Artifact
 - PTV_WB
 - PTV_WBopt
 - Reye
 - RLacrimal
 - RLens
 - Roptic






Use Default Models	Particle Type	Calculation Type	Calculation Model	Status	Algorithm	Calculation Options
Clear All Selections	Photon	Volume Dose	AAA_15606	OK		
		Portal Dose	AAA_15606	No configured beam data (Eclipse CAP 6X PORTALDOSE)	Anisotropic Analytical Algorithm (Version 15.6.06)	Edit
		DVH Estimation	DVH Estimation Algorithm [15.6.06]	OK	DVH Estimation Algorithm (Version 15.6.06)	Edit
		Beam Angle Optimization	PGO_15161	OK	Plan Geometry Optimizer (Version 15.1.61)	Edit
		IMRT Optimization	PO_15606	OK		
		VMAT Optimization	PO_15606	OK	Photon Optimizer (Version 15.6.06)	Edit
		Irregular Surface Compensator	PO_15606	Not supported for this plan		

HCSWB

- Etapa 1
 - IMRT
 - VMAT



IMRT

- HCSWB V2.0
 - Registered Images
 - HCSWB V2.0
 - User Origin
 - Reference Points
 - PTV_WBopt
 - Dose
 - Fields
 - Isocenter Group 1
 - Field 1
 - Fluence
 - Field 2
 - Fluence
 - Field 3
 - Fluence
 - Field 4
 - Fluence
 - Field 5
 - Fluence
 - Field 6
 - Fluence
 - Field 7
 - Fluence
 - Field 8
 - Fluence



Use Default Models	Particle Type	Calculation Type	Calculation Model	Status	Algorithm	Calculation Options	
Clear All Selections	Photon	Volume Dose	AAA_15606	OK	Anisotropic Analytical Algorithm (Version 15.6.06)	Edit	
		Portal Dose	AAA_15606	No configured beam data (Eclipse CAP 6X PORTALDOSE)		DVH Estimation Algorithm (Version 15.6.06)	Edit
		DVH Estimation	DVH Estimation Algorithm [15.6.06]	OK		Plan Geometry Optimizer (Version 15.1.61)	Edit
		Beam Angle Optimization	PGO_15161	OK		Photon Optimizer (Version 15.6.06)	Edit
		IMRT Optimization	PO_15606	OK			
		VMAT Optimization	PO_15606	OK			
	Irregular Surface Compensator	PO_15606	Not supported for this plan				

Plan Information							
ID/Type	cm ³	Vol [%]	Dose[Gy]	Actual Dose[Gy]	Priority	gEUD a	
<input checked="" type="checkbox"/> PTV_WBopt	1192.7						
Upper	0.0	0.0	3060.00	3278.44	100		x
Lower	1180.8	99.0	2940.00	2654.27	100		x
<input checked="" type="checkbox"/> Eyes	13.8						
Upper	0.0	0.0	700.00	1171.19	50		x
<input checked="" type="checkbox"/> Hippocampus	4.4						
Upper	0.0	0.0	1100.00	1993.29	100		x
Upper	1.7	40.0	900.00	1194.69	100		x
<input checked="" type="checkbox"/> Leye	6.8						
Upper	0.0	0.0	700.00	1179.11	50		x
<input checked="" type="checkbox"/> LLens	0.1						
Upper	0.0	0.0	500.00	553.56	50		x
<input checked="" type="checkbox"/> Reye	7.0						
Upper	0.0	0.0	700.00	967.51	50		x
<input checked="" type="checkbox"/> RLens	0.2						
Upper	0.0	0.0	500.00	629.83	50		x
<input checked="" type="checkbox"/> _Brain&BODY	958.6						
<input checked="" type="checkbox"/> _Brainstem#H	19.5						
<input checked="" type="checkbox"/> _Eyes&Body	916.4						
<input checked="" type="checkbox"/> BODY	6255.0						
Normal Tissue Objective							100/Manual
Base Dose Plan							None
Settings							1000/6000s/Normal (2.5 mm)



Progress | **Plan Objectives**

3D Dose Max: 3322.51 Gy
 3D MAX for PTVWBopt: 3278.44 Gy
 3D MEAN for PTVWBopt: 3030.06 Gy
 3D MIN for PTVWBopt: 1470.53 Gy



Elapsed Time
 Iteration

Fluence
 Intermediate Dose

Open Log...

Automatic Optimization Mode
 Automatic Intermediate Dose

Plan Information							
ID/Type	cm ³	Vol [%]	Dose[Gy]	Actual Dose[Gy]	Priority	gEUD a	
<input checked="" type="checkbox"/> PTV_WBopt	1192.7						
Upper	0.0	0.0	3060.00	3278.44	100		x
Lower	1180.8	99.0	2940.00	2654.27	100		x
<input checked="" type="checkbox"/> Eyes	13.8						
Upper	0.0	0.0	700.00	1171.19	50		x
<input checked="" type="checkbox"/> Hippocampus	4.4						
Upper	0.0	0.0	1100.00	1993.29	100		x
Upper	1.7	40.0	900.00	1194.69	100		x
<input checked="" type="checkbox"/> Leye	6.8						
Upper	0.0	0.0	700.00	1179.11	50		x
<input checked="" type="checkbox"/> LLens	0.1						
Upper	0.0	0.0	500.00	553.56	50		x
<input checked="" type="checkbox"/> Reye	7.0						
Upper	0.0	0.0	700.00	967.51	50		x
<input checked="" type="checkbox"/> RLens	0.2						
Upper	0.0	0.0	500.00	629.83	50		x
<input checked="" type="checkbox"/> _Brain&BODY	958.6						
<input checked="" type="checkbox"/> _Brainstem#H	19.5						
<input checked="" type="checkbox"/> _Eyes&Body	916.4						
<input checked="" type="checkbox"/> BODY	6255.0						
Normal Tissue Objective							100/Manual
Base Dose Plan							None
Settings							1000/6000s/Normal (2.5 mm)





Progress	Plan Objectives
	3D Dose Max 3322.51 Gy
	3D MAX for PTVWBopt 3278.44 Gy
	3D MEAN for PTVWBopt 3030.06 Gy
	3D MIN for PTVWBopt 1470.53 Gy
	Elapsed Time
	Iteration
<input checked="" type="checkbox"/> Fluence	
<input checked="" type="checkbox"/> Intermediate Dose	

Open Log...

Automatic Optimization Mode
 Start IMRT Optimization
 Intermediate Dose

OK Cancel


Plan Information						
ID/Type	cm ³	Vol [%]	Dose[Gy]	Actual Dose[Gy]	Priority	gEUD a
<input checked="" type="checkbox"/> PTV_WBopt	1192.7					
Upper	0.0	0.0	3060.00	3359.66	100	x
Lower	1180.8	99.0	2940.00	2480.55	100	x
<input checked="" type="checkbox"/> Eyes	13.8					
Upper	0.0	0.0	700.00	1232.36	50	x
<input checked="" type="checkbox"/> Hippocampus	4.4					
Upper	0.0	0.0	1000.00	1242.64	120	x
Upper	1.7	40.0	900.00	914.49	100	x
<input checked="" type="checkbox"/> Leye	6.8					
Upper	0.0	0.0	700.00	1601.86	50	x
<input checked="" type="checkbox"/> LLens	0.1					
Upper	0.0	0.0	500.00	511.87	50	x
<input checked="" type="checkbox"/> Reye	7.0					
Upper	0.0	0.0	700.00	982.59	50	x
<input checked="" type="checkbox"/> RLens	0.2					
Upper	0.0	0.0	500.00	545.61	50	x
<input checked="" type="checkbox"/> _Brain&BODY	958.6					
<input checked="" type="checkbox"/> _Brainstem#H	19.5					
<input checked="" type="checkbox"/> _Eyes&Body	916.4					
<input checked="" type="checkbox"/> BODY	6255.0					
Normal Tissue Objective					100/Manual	
Base Dose Plan					None	
Settings					1000/6000s/Normal (2.5 mm)	



Progress | Plan Objectives

3D Dose Max: 3362.14 Gy
 3D MAX for PTVWBopt: 3359.66 Gy
 3D MEAN for PTVWBopt: 2984.11 Gy
 3D MIN for PTVWBopt: 1073.91 Gy
 Elapsed Time: 65 s
 Iteration: 174

Fluence
 Intermediate Dose



Organ	Value
PTV_WBopt	3362.14
BODY	3359.66
Hippocampus	2984.11
Reye	1073.91
Leye	1601.86
Eyes	1232.36
Roptic	982.59
RLens	545.61

Open Log...



Automatic Optimization Mode
 Stop IMRT Optimization
 Intermediate Dose
 Optimizing...
 Converged
 OK
 Cancel

HCSWB

- Etapa 1
 - IMRT
 - VMAT

HCSWB V2.0

- Registered Images
- HCSWB V2.0
 - _Brain&BODY
 - _Brainstem#Hi
 - _Eyes&Body
 - BODY
 - Brain
 - Brainstem
 - Chiasm
 - Cord
 - Eyes
 - Hippo+05
 - Hippocampus
 - Leye
 - LLacrimonal
 - LLens
 - Loptic
 - NS_Artifact
 - PTV_WB
 - PTV_WBopt
 - Reye
 - RLacrimonal
 - RLens
 - Roptic
 - User Origin



Calculation Progress

ID	#	Servant ID	Progress	State
Field 1	Volume Dose	p4988@calc-h48bd71	100%	Completed
Field 2	Volume Dose	p15240@calc-h48bd71	100%	Completed
Field 3	Volume Dose	p7760@calc-h48bd71	100%	Completed
Field 4	Volume Dose	p27860@calc-h48bd71	100%	Completed
Field 5	Volume Dose	p25508@calc-h48bd71	0%	Calculating
Field 6	Volume Dose	p25852@calc-h48bd71	0%	Calculating
Field 7	Volume Dose	p13980@calc-h48bd71	0%	Calculating
Field 8	Volume Dose	p10176@calc-h48bd71	0%	Calculating
Field 9	Volume Dose		0%	Waiting...

44%

Close after successful calculation

Abort Close





Show DVH	Structure	Approval Status	Plan	Course	Volume [cm ³]	Dose Cover. [%]	Sampling Cover. [%]	Min Dose [?]	Max Dose [?]	Mean Dose [?]
<input type="checkbox"/>	RLacrimonal	Unapproved	IMRT	Etapa1						
<input type="checkbox"/>	LLacrimonal	Unapproved	IMRT	Etapa1						
<input type="checkbox"/>	Brain	Unapproved	IMRT	Etapa1						
<input type="checkbox"/>	Cord	Unapproved	IMRT	Etapa1						
<input checked="" type="checkbox"/>	Hippocampus	Unapproved	IMRT	Etapa1						
<input type="checkbox"/>	Hippo+05	Unapproved	IMRT	Etapa1						
<input type="checkbox"/>	BODY	Unapproved	IMRT	Etapa1						

HCSWB

- Etapa1
 - IMRT
 - VMAT

HCSWB V2.0





- Registered Images
- HCSWB V2.0
 - _Brain&BODY
 - _Brainstem#Hi
 - _Eyes&Body
 - BODY
 - Brain
 - Brainstem
 - Chiasm
 - Cord
 - Eyes
 - Hippo+05
 - Hippocampus
 - Leye
 - LLacrimal
 - LLens
 - Loptic
 - NS_Artifact
 - PTV_WB
 - PTV_WBopt
 - Reye
 - RLacrimal
 - RLens
 - Roptic
 - User Origin



Show DVH	Structure	Approval Status	Plan	Course	Volume [cm ³]	Dose Cover. [%]	Sampling Cover. [%]	Min Dose [Gy]	Max Dose [Gy]	Mean Dose [Gy]
<input type="checkbox"/>	PTV_WB	Unapproved	IMRT	Etapa1						
<input type="checkbox"/>	Brainstem	Unapproved	IMRT	Etapa1						
<input type="checkbox"/>	Chiasm	Unapproved	IMRT	Etapa1						
<input type="checkbox"/>	Roptic	Unapproved	IMRT	Etapa1						
<input type="checkbox"/>	Loptic	Unapproved	IMRT	Etapa1						
<input checked="" type="checkbox"/>	Reye	Unapproved	IMRT	Etapa1		7.0	100.0	100.1	471.881	871.212
<input checked="" type="checkbox"/>	RLens	Unapproved	IMRT	Etapa1		0.2	100.0	99.1	488.841	558.938
<input checked="" type="checkbox"/>	Leye	Unapproved	IMRT	Etapa1		6.8	100.0	99.7	458.041	1211.292

HCSWB

- Etapa 1
 - IMRT
 - VMAT
- HCSWB V2.0
 - Registered Images
 - HCSWB V2.0
 - User Origin
 - Reference Points
 - PTV_WBopt
 - Dose
 - Fields
 - Isocenter Group 1
 - Field 1
 - Fluence
 - Field 2
 - Fluence
 - Field 3
 - Fluence
 - Field 4
 - Fluence
 - Field 5
 - Fluence
 - Field 6
 - Fluence
 - Field 7
 - Fluence
 - Field 8
 - Fluence
 - Field 9



Show DVH	Structure	Approval Status	Plan	Course	Volume [cm ³]	Dose Cover. [%]	Sampling Cover. [%]	Min Dose [Gy]	Max Dose [Gy]	Mean Dose [Gy]	
<input type="checkbox"/>	PTV_WB	Unapproved	IMRT	Etapa1							
<input type="checkbox"/>	Brainstem	Unapproved	IMRT	Etapa1							
<input type="checkbox"/>	Chiasm	Unapproved	IMRT	Etapa1							
<input type="checkbox"/>	Optic	Unapproved	IMRT	Etapa1							
<input type="checkbox"/>	Loptic	Unapproved	IMRT	Etapa1							
<input checked="" type="checkbox"/>	Reye	Unapproved	IMRT	Etapa1		7.0	100.0	100.1	471.881	871.212	625.980
<input checked="" type="checkbox"/>	RLens	Unapproved	IMRT	Etapa1		0.2	100.0	99.1	488.841	558.938	512.807
<input checked="" type="checkbox"/>	Leye	Unapproved	IMRT	Etapa1		6.8	100.0	99.7	458.041	1211.292	630.450